



© 2015 Roche Diabetes Care

ACCU-CHEK, ACCU-CHEK MOBILE, ACCU-CHEK SMART PIX, ACCU-CHEK 360° and FASTCLIX are trademarks of Roche.

CONTINUA, the CONTINUA logos and CONTINUA CERTIFIED are trademarks, service marks, or certification marks of the Continua Health Alliance. CONTINUA is a registered trademark in some, but not all countries in which this product is distributed.

The USB-IF Logos are trademarks of Universal Serial Bus Implementers Forum, Inc.

Roche Diabetes Care GmbH
Sandhofer Strasse 116
68305 Mannheim, Germany

www.accu-chek.com

0743887001(01) – 2015-05 M_Version02_Rev3

ACCU-CHEK®

ACCU-CHEK® Mobile



User's Manual and Quick Start Guide

Blood Glucose Monitoring System





ACCU-CHEK®


IMPORTANT INFORMATION FOR PERFORMING A BLOOD GLUCOSE TEST


An incorrectly performed blood glucose test may lead to incorrect test results which can cause the wrong therapy recommendation to be made and thus produce serious adverse health effects. Therefore, **follow** these instructions carefully to perform a blood glucose test correctly.

Do

1  Wash your hands with warm water and soap and rinse well. Dry your hands thoroughly with a clean towel before obtaining blood.


2  Always use a small, fresh and well-formed blood drop.


3  Touch the test area gently and steadily when applying the blood drop. Apply the blood drop to the centre of the test area.

4  **IMMEDIATELY!** Apply the blood drop to the test area immediately after you have obtained the blood drop. Remove your finger from the test area when the beep tone sounds and *Test in progress* is displayed. **Note:** You can only hear the beep tones when they are turned on. See *Setting tones* in the User's Manual.

Do not

 Do **not** perform a blood glucose test if your fingers are unwashed, soiled or sticky (for example, due to food or drink residue).

 Do **not** perform a blood glucose test with a smeared or spread blood drop. Do **not** smear the blood drop on the test area.

 Do **not** press the test area and the finger together. Avoid trembling or shaky movements.

 **DO NOT WAIT!** Do **not** wait before applying the blood drop to the test area. Do **not** continue to hold your finger on the test area after the beep tone sounds and *Test in progress* is displayed.

Why

An incorrectly performed blood glucose test may lead to falsely elevated blood glucose results and thereby delivery of an inappropriately high insulin amount. Falsely elevated test results are caused by the following:

- Fingers are unwashed, soiled or sticky.
- If the blood drop spreads on the finger or if the blood drop becomes smeared on the test area, the risk of contamination of the test area increases (for example, due to food or drink residue on the finger).
- The test area is pressed against the finger.
- Trembling or shaky movements during the blood glucose test.
- The blood glucose test is not performed immediately or the finger stays on the test area for too long. Pay attention to the beep tone. The beep tone helps you to obtain a reliable test result.

ACCU-CHEK® Mobile



Quick Start Guide

Start Here

ACCU-CHEK®



- 1** Release button (priming and pricking)
- 2** Power and enter button
- 3** Down/Up buttons
- 4** Window showing number of available lancets
- 5** Lever for loading a new lancet
- 6** Rotatable cap for setting the penetration depth
- 7** Tip cover, closed (below right: open)
- 8** Test area (at the tip of the cassette ready for a test)
- 9** Lancet drum
- 10** Slide button to release the finger pricker
- 11** Slide button to open the cassette compartment cover
- 12** Tip cover
- 13** Test cassette



The Accu-Chek Mobile meter and Accu-Chek FastClix finger pricker are intended for patient self-monitoring by an individual person only.

They must not be used to test blood glucose from more than one person as they do not incorporate any features to guard against cross-infection. The meter and the finger pricker are not to be shared between family members or used by healthcare professionals to obtain blood or test blood glucose from more than one person.

This Quick Start Guide does not replace the detailed User's Manual for your Accu-Chek Mobile blood glucose monitoring system. Be sure to comply with the safety information in the User's Manual and in the package insert of the test cassette.

1. Starting to use the blood glucose monitoring system



Open the test cassette box at the perforation. Take out the plastic container.



Remove the new test cassette from the plastic container.



Open the tip cover.



Push the slide button for the cassette compartment cover upwards in the direction of the arrow.



Insert the test cassette in the meter. Press the cassette compartment cover closed. Close the tip cover.



Pull the protective strip of the batteries backwards out of the meter. Remove the protective film from the display.



Press for about 2 seconds until the meter turns on.



Use to select a language (highlighted in yellow). Press .



Press .

2. Preparing the finger pricker



Remove the cap from the finger pricker.



Insert the new lancet drum, white end first.




Do **not** remove the lancet drum until all 6 lancets have been used.



The lancet drum is properly inserted when it is completely inside the finger pricker and only the front edge protrudes.



Place the cap back on again.



The cap is on correctly when it is pushed back as far as it will go, and you hear and feel it click into place.



Rotate the cap until the desired penetration depth lines up with the indicator.



Loading a new lancet: Prepare the finger pricker for obtaining blood again.

Tip: When obtaining blood, start with a medium penetration depth, such as 3.

Push the lever in direction **A** and then back again in direction **B**.

3. Performing a blood glucose test

Before you perform a blood glucose test: Observe the important instructions on performing a blood glucose test.



Wash your hands with warm water and soap and rinse well.



Dry your hands thoroughly with a clean towel before you obtain blood.



Open the tip cover.




Press the finger pricker firmly against the selected puncture site.



Press the release button all the way down.



Touch the centre of the test area gently and steadily with the blood drop.



Take your finger away from the test area as soon as *Test in progress* appears on the screen.



Testing is complete after approximately 5 seconds. Read the test result and close the tip cover.

Customer Support and Service Centre

Australia
Accu-Chek Enquiry Line: 1800 251 816
Pump Support: 1800 633 457
www.accu-chek.com.au

Singapore
Accu-Chek ExtraCare line: 6272 9200
www.accu-chek.com.sg

United Kingdom
Roche Diabetes Care Limited
Charles Avenue, Burgess Hill
West Sussex, RH15 9RY, United Kingdom
Accu-Chek Customer Careline ¹⁾
UK Freephone number: 0800 701 000
ROI Freephone number: 1 800 709 600
¹⁾ calls may be recorded for training purposes
www.accu-chek.co.uk
www.accu-chek.ie

Last update: 2015-02



© 2015 Roche Diabetes Care

ACCU-CHEK, ACCU-CHEK MOBILE and FASTCLIX are trademarks of Roche.

Roche Diabetes Care GmbH
Sandhofer Strasse 116
68305 Mannheim, Germany
www.accu-chek.com

ACCU-CHEK® Mobile

Roche

User's Manual

Blood Glucose Monitoring System



ACCU-CHEK®

These instructions for use feature the following 3 symbols:



This symbol indicates a **possible risk of injury or of damage to your own health or to the health of others.**



Keep the blood glucose monitoring system and all its components away from children under the age of 3 years. There is a risk of suffocation if small parts (e.g. covers, caps or similar objects) are swallowed.



This symbol draws attention to actions that could result in **damage to the meter, the test cassette or the finger pricker.**



This symbol draws your attention to **important information.**

Intended use

Accu-Chek Mobile blood glucose meter

Meter for quantitative determination of blood glucose values in fresh capillary blood using Accu-Chek Mobile test cassettes.

Suitable for self-testing only.



- **The Accu-Chek Mobile meter and Accu-Chek FastClix finger pricker are intended for patient self-monitoring by an individual person only.**

They must not be used to test blood glucose from more than one person as they do not incorporate any features to guard against cross-infection. The meter and the finger pricker are not to be shared between family members or used by healthcare professionals to obtain blood or test blood glucose from more than one person. This meter and finger pricker are therefore not suitable for professional use in healthcare facilities and institutions.

- Visually impaired people must be assisted by a sighted person when performing a blood glucose test.

Accu-Chek FastClix finger pricker

Finger pricker with adjustable penetration depth for obtaining capillary blood from the fingertip with Accu-Chek FastClix lancet drums and, when a special AST cap is used, also from alternative sites (AST = alternative site testing), such as the palm.



- **The Accu-Chek FastClix finger pricker is intended for personal use only!**

It may only be used by **one and the same** person for obtaining blood.

There is a risk of infections being transmitted if the finger pricker is used by other people, even by family members, or if healthcare professionals use this finger pricker to obtain blood from different people. This finger pricker is therefore not suitable for professional use in healthcare facilities and institutions.

Contents

Intended use	3
The major features	5
About this User's Manual	6
1 Getting to know your meter and finger pricker	7
2 Steps before testing	14
3 Preparing the finger pricker	21
4 Performing a blood glucose test	27
5 Selecting settings	45
6 Using the meter as a diary	64
7 Analysing test results on the PC	70
8 Acoustic mode	89
9 Checking the meter	94
10 Tools menu	101
11 Cleaning the blood glucose monitoring system	107
12 Changing the batteries	114
13 Testing and storage conditions	117
14 Messages and problems	120
15 Discarding the blood glucose monitoring system	129
16 Technical data	130
17 System components	133
18 Customer Support and Service Centre	134
19 Index	135
20 Explanation of symbols	138
21 Appendix	140

The major features

- **Test cassette instead of test strips**
50 test areas on a continuous tape
- **Test**
Test starts by opening the tip cover
- **Measuring time**
About 5 seconds for a test, depending on the blood glucose concentration
- **Docked finger pricker**
Use the finger pricker in its docked or undocked state
- **Finger pricker with lancet supply**
6 sterile lancets in a lancet drum
- **Automatic coding**
Meter is automatically coded
- **Control over amount of blood**
Meter detects the amount of blood that is necessary
- **Luminous display**
Yellow display on black background
- **Text-supported operation**
Meter takes you through all operation steps
- **Menu-driven operation**
Menu-driven settings and functions
- **Reminder**
A total of 7 reminders and 4 test reminders
- **Target range for test results**
Personal target range for blood glucose values
- **Flagging of results**
Flagging results with symbols
- **Memory**
2,000 memory locations available
- **Data analysis on a PC**
Display and analysis of data on a PC
- **Data transfer**
USB port for data transfer
- **Error messages**
Displays an error code and a message

About this User's Manual


Read this User's Manual carefully and completely before testing blood glucose for the first time. If you have any questions, contact the customer support and service centre (see page 134).

These instructions for use provide you with the information you need to operate and care for your meter and finger pricker, and for troubleshooting. Be sure to operate the meter and the finger pricker correctly as well as to observe the operating instructions.

Note: All dates, times or results displayed on the screens in this User's Manual are intended only as examples. Results are displayed in the unit *mmol/L*, dates in the format *Day.Month.Year* and times in *24-hour format*.

You can only hear the beep tones of the meter when they are turned on and the volume level is not set to *1 (Mute)*. This User's Manual presumes that the beep tones are turned on (see *Setting tones* page 49).

All instructions appear as shown in the example below.




Getting to know your meter and finger pricker

Overview of meter



- 1 Battery door (above right: open)
- 2 Batteries (see page 114 and page 130 for information on possible battery types)
- 3 Display
- 4 Finger pricker (docked)
- 5 Power and enter button
- 6 Down/Up buttons
- 7 Tip cover, closed (below right: open)
- 8 Test area (at the tip of the cassette, ready for a test)
- 9 USB port – interface to be used for data analysis on a computer (below left: open USB port)
- 10 Type plate
- 11 Slide button to release the finger pricker
- 12 Slide button to open the cassette compartment cover
- 13 Cassette compartment cover (below right: open)
- 14 Test cassette (in place in the meter)

Overview of finger pricker



- 15 Release button (priming and pricking)
- 16 Lever for loading a new lancet
- 17 Lancet drum containing 6 lancets
- 18 Numerical values for penetration depths (11 settings)
- 19 Rotatable cap for setting the penetration depth
- 20 Pin-hole opening for lancet
- 21 Penetration depth indicator
- 22 Window showing number of available lancets


Using the buttons and menus

This chapter tells you how to use the buttons to operate the meter and navigate through the menus, how the menus are structured and what the different symbols in them mean.

Meter buttons


Button	Name	Symbol	Function
	Power button, enter button		This button allows you to: <ul style="list-style-type: none">• Turn the meter on or off; to do so, press and hold the button (for approx. 2 seconds)• Execute the command displayed on the screen above the button (press and release the button only) Example: The Select command selects the option highlighted in the menu.
	Down button		This button 1 allows you to: <ul style="list-style-type: none">• Navigate down in a menu or list• Decrease numerical values
	Up button		This button 2 allows you to: <ul style="list-style-type: none">• Navigate up in a menu or list• Increase numerical values

Menu structure of the meter




- 1 Menu
- 2 Title of a screen or menu (for example, *Settings*)
- 3 The symbol means that there are more options listed above the first visible option.
- 4 The command displayed here (for example, *Select*) will be executed when you press the  button. The command that is shown depends on the current screen.
- 5 The symbol means that there are more options listed below the last visible option.
- 6 A selected option is highlighted with a yellow bar. If you press the  button, this option will be selected.
- 7 The  symbol indicates that the function (for example, *Reminders*) is on.
- 8 The  symbol indicates that the function (for example, *Tones*) is off.
- 9 Unselected option in the menu.
- 10 The options in the main menu and the *Language* option in the *Settings* menu have a symbol on their left-hand side to identify the menu item should the language accidentally change to one that you do not understand ( for *Language* for example).

Operation



Pressing and holding the button  (for approx. 2 seconds) **1** turns the meter on.

After the display check, the meter first displays the number of tests still available on the test cassette and then opens the main menu.




Select 2 (press and release the ) selects the highlighted *Turn off* command (.

Or


Press and hold the  button until all bars on the screen are filled (approx. 2 seconds) **3**.

The meter turns off after displaying the number of tests still available.




Select 4 (press and release the ) selects the highlighted *Time/Date* option from the *Settings* menu. The *Time/Date* menu opens **5**.

Returning to a higher-level menu




Back **6** (press and release the button) opens the higher-level menu **7** of the currently displayed menu.

Confirming settings or operations




Confirm (**8**) is highlighted.
OK **9** (press and release the button) confirms the setting you made.

Cancelling settings or operations




(*Cancel*) is highlighted.
OK **10** (press and release the button) cancels the setting you made.
Cancel **11** (press and release the button) cancels the operation (data transfer).

Navigating between input fields



Continue (press and release the  button) switches from month  to year .

Changing numerical values




Pressing the  button  increases the numerical value.

Pressing the  button  decreases the numerical value.

Pressing the button once increases or decreases the number by 1. If you press and hold the button, the number continues to count until you release the button.

Turning functions on or off



A tick   in front of an option means that the function is on.

A checkbox   in front of an option means that the function is off.

Select (press and release the  button) sets (= on) or removes (= off) the tick.

Steps before testing

Checking the unit of measurement

Blood glucose results can be displayed in two different units of measurement (mg/dL or mmol/L). Consequently, there are two different versions of the same meter. Check that your meter displays the unit of measurement you are accustomed to. You can find the unit of measurement **1** that your meter displays on the type plate on the back of your meter. If you do not know which unit of measurement is correct for you, ask your healthcare professional.



The unit of measurement that your meter displays cannot be changed. If the wrong unit of measurement is printed on the type plate, contact the customer support and service centre (see page 134). Using the wrong unit of measurement may cause misinterpretation of your test results and can cause the wrong therapy recommendation to be made, and thus produce serious adverse health effects.

Type plate



Inserting the first test cassette

Before using your new meter for the first time, you must insert a test cassette.

Insert the very first test cassette in the meter before you remove the protective strip from the batteries and start using the meter.



- Read the test cassette package insert. It contains further important information, for example, on storage and possible causes of incorrect test results.
- If the plastic container or the foil cover of the test cassette is damaged, you must not use the test cassette. Otherwise, there is a risk that you might obtain incorrect test results. Incorrect test results can cause the wrong therapy recommendation to be made and thus produce serious adverse health effects.
- Only open the plastic container when you want to insert the test cassette in the meter. The unopened plastic container protects the test cassette against damage and moisture.



A table is printed on the test cassette box which shows the permitted results for control tests (checking the meter with glucose control solution). The meter automatically checks whether the result of a control test is correct (see page 98). If you want to check the control result yourself as well, you can do this with this table. Keep the test cassette box in case you need it for this purpose. Note that the table only applies to test cassettes in this box. For test cassettes from other boxes, other tables apply.

2


Steps before testing

1




Open the test cassette box at the perforation. Take out the plastic container.

2



Remove the test cassette from the plastic container.

3



Open the tip cover.

4



Turn the meter over so that the back is facing upwards.


5




Push the slide button for the cassette compartment cover upwards in the direction of the arrow.

The cassette compartment cover opens.

6



Insert the test cassette in the meter with the silver side facing upwards.




Close the cassette compartment cover.



Press the cassette compartment cover closed.

The cassette compartment cover must close with an audible **CLICK**.



Close the tip cover.

Validity of the test cassette

The validity of the test cassette depends on the use by period and the use by date.

Use by period: The period in which the test cassette must be used up after the foil cover of the plastic container was opened. The use by period in days is printed on the box and in the package insert of the test cassette next to the  symbol.

Use by date: Date until which a test cassette sealed in the plastic container is valid. The use by date is shown on the box of the test cassette or foil cover next to the symbol .

If either of the two dates – use by period or use by date – is exceeded, you will not be able to perform any more tests with this test cassette.

If you start a test and the validity of the test cassette will soon expire or has already expired, a message appears to inform you.

The first message appears 10 days before the validity expires, the others follow 5, 2 and 1 day(s) before expiry (see *Error messages* page 124).


If the validity of the test cassette has expired, a message appears (see *Error messages* page 124).

Removing protective film and strip

The meter display is covered with a protective film and the meter's battery contacts with a protective strip.




Remove the protective film from the display.



Pull the protective strip of the batteries backwards out of the meter.

Setting the language


The meter automatically prompts you to set the language the first time you turn it on. When you turn the meter on for the first time, the language selection menu is automatically displayed. You can either select the default language or replace it with a different language.



Use **▼ ▲** to select a language.


Press **■** to select the desired language.


The meter displays the language you have selected.



Press **■** to confirm the selected language.

The meter then runs a display check.






If there are irregularities in the checkerboard pattern, results might not be displayed correctly. In this case, contact the customer support and service centre (see page 134).

Check if there are any irregularities in the checkerboard pattern ① (see *Performing a display check* page 101).

The meter automatically exits the display check after about 2 seconds.



The meter displays the number of tests in the test cassette that are still available.

Afterwards the main menu is shown.

Now you can decide what you want to do (for example, turn the meter off, perform a blood glucose test or make settings).

Preparing the finger pricker

You can use the finger pricker to obtain blood from a fingertip. You can set the penetration depth to 11 different levels to suit the individual texture of your skin.

You insert a new lancet drum with 6 sterile lancets into the finger pricker. You can obtain blood with a sterile lancet 6 times before the lancet drum needs changing.

You can use the finger pricker either docked on to the meter or separately from the meter.

You can also use the finger pricker to obtain blood from other (alternative) sites. For this purpose, you need a special AST cap, which enables you to obtain capillary blood from alternative sites.

If you would like to order an AST cap, contact the customer support and service centre (see page 134).



- **The Accu-Chek Mobile meter and Accu-Chek FastClix finger pricker are intended for patient self-monitoring by an individual person only.**

They must not be used to test blood glucose from more than one person as they do not incorporate any features to guard against cross-infection. The meter and the finger pricker are not to be shared between family members or used by healthcare professionals to obtain blood or test blood glucose from more than one person. This meter and finger pricker are therefore not suitable for professional use in healthcare facilities and institutions.


- Wear and tear of the materials can lead to malfunction of the device. In extreme cases a lancet can protrude from the cap so that injury cannot be completely ruled out. You should therefore handle the finger pricker with particular care once a lancet drum has been inserted.



- Only use the Accu-Chek FastClix finger pricker with Accu-Chek FastClix lancet drums. Using any other lancets or lancet drums may severely damage the finger pricker or impair its function.
- A used Accu-Chek FastClix lancet drum has a built-in locking mechanism. You cannot reinsert a lancet drum that has already been removed. Used lancet drums must not be reused.

3

Preparing the finger pricker



Take a new lancet drum ①.

You can recognise used lancet drums by the red stripe ②.



Remove the cap from the finger pricker.



You must not insert the lancet drum into the finger pricker and simultaneously press the release button or hold the finger pricker with the release button resting on a surface such as a table top. This could release a lancet and inadvertently cause injury.



Insert the new lancet drum, white end first.

The lancet drum is properly inserted when it is completely inside the finger pricker and only the front edge protrudes.



Place the cap back on again.

The cap is on correctly when it is pushed back as far as it will go, and you hear and feel it click into place.



The cap can only be easily placed on up to the stop position when the lancet drum is completely inside the finger pricker. Do not use force to push the cap on. This could damage the cap and finger pricker. Instead, correct the position of the lancet drum.

Setting the penetration depth

You can set the penetration depth of the lancet to 11 different levels. The levels are divided into 6 half-settings and 5 whole settings (0.5–5.5, 0.5 is the lowest, 5.5 is the highest). Set the penetration depth suitable for you. This allows virtually pain-free collection of blood and control over the amount of blood needed. The penetration depth indicator on the finger pricker (silver square) indicates the current penetration depth setting. The higher the number, the greater the penetration depth. The half-settings are located between the numbers.

If you have no experience with using this finger pricker, we recommend a medium penetration depth setting such as 3.

Settings



1 = penetration depth indicator

2 = half-setting



Rotate the cap until the desired penetration depth is level with the indicator.

Loading a new lancet

When you have inserted a new lancet drum, a new lancet is automatically ready for use.

If you have already used the finger pricker to obtain blood, you will need a new lancet the next time you obtain blood. The following steps show you how to load a new lancet:



Push the lever 1 in direction A and then back again in direction B.



The window 2 on the side of the finger pricker now shows one less available lancet (for example, 5 instead of previously 6).



If the number 1 appears in the window showing the number of available lancets 2, the lever 1 will be blocked by a locking mechanism. Do not force the lever beyond this locking mechanism as this will damage the finger pricker. Instead, replace the lancet drum with a new one.

Replacing the lancet drum


When you have used the sixth and last lancet, the number 1 appears in the window on the finger pricker. Replace the lancet drum with a new one.




A used lancet drum has a built-in locking mechanism and cannot be reused.



Used lancet drums can be disposed of in household waste if no other regulations apply locally.



Remove the cap from the finger pricker.



Pull the lancet drum out of the finger pricker.



Insert a new lancet drum, white end first.

Place the cap back on again.

Keep in mind the information on page 22.


3

Preparing the finger pricker

Undocking the finger pricker



Push the slide button for the finger pricker in the direction of the arrow as far as it will go. Hold the slide button in that position.




Slide the finger pricker out in the direction of the arrow.

Docking the finger pricker



When docking the finger pricker, the tracks of the finger pricker **1** and those of the meter **2** must interlock.



Slide the finger pricker along the side of the meter in the direction of the arrow until it slots into place.

Performing a blood glucose test

Once you have inserted a lancet drum into the finger pricker and set the penetration depth, you can start to perform a blood glucose test.



Important information for performing a blood glucose test





Performing a blood glucose test incorrectly may lead to incorrect test results which can cause the wrong therapy recommendation to be made and thus produce serious adverse health effects. Therefore, follow these instructions carefully to perform a blood glucose test correctly.



Visually impaired people must be assisted by a sighted person when applying a blood drop.
If you are using the acoustic mode: Pay attention to the beep tones. Read the *Acoustic mode* chapter carefully (see page 89).




When the meter prompts you to apply a blood drop, you have about 2 minutes to apply blood to the test area. If you do not apply blood during this time, the meter turns itself off and the test area will be lost.

		Do	
1			Wash your hands with warm water and soap and rinse well. Dry your hands thoroughly with a clean towel before obtaining blood.
2			Always use a small, fresh and well-formed blood drop.
3			Touch the test area gently and steadily when applying the blood drop. Apply the blood drop to the centre of the test area.
4			Apply the blood drop to the test area immediately after you have obtained the blood drop. Remove your finger from the test area when the beep tone sounds and <i>Test in progress</i> is displayed. Note: You can only hear the beep tones when they are turned on. See <i>Setting tones</i> page 49.


Do not



Do **not** perform a blood glucose test if your fingers are unwashed, soiled or sticky (for example, due to food or drink residue).



Do **not** perform a blood glucose test with a smeared or spread blood drop. Do **not** smear the blood drop on the test area.



Do **not** press the test area and the finger together. Avoid trembling or shaky movements.



DO NOT WAIT!

Do **not** wait before applying the blood drop to the test area. Do **not** continue to hold your finger on the test area after the beep tone sounds and *Test in progress* is displayed.

Why

An incorrectly performed blood glucose test can lead to falsely elevated blood glucose results and thereby delivery of an inappropriately high insulin amount. Falsely elevated test results are caused by the following:

- Fingers are unwashed, soiled or sticky.
- If the blood drop spreads on the finger or if the blood drop becomes smeared on the test area, the risk of contamination of the test area increases (for example, due to food or drink residue on the finger).
- The test area is pressed against the finger.
- Trembling or shaky movements during the blood glucose test.
- The blood glucose test is not performed immediately or the finger stays on the test area for too long. Pay attention to the beep tone. The beep tone helps you to obtain a reliable test result.

Starting a blood glucose test

There are 2 ways to start a blood glucose test: By opening the tip cover (A) or from the *Main menu* (B).


A – Starting a blood glucose test by opening the tip cover:



Open the tip cover.

If the meter was turned off, it now turns on.

A beep tone sounds and the meter runs a display check.



Check that there are no irregularities in the checkerboard pattern **1** (see also *Performing a display check* page 101).



The number of available tests **2** and the prompt to wash your hands **3** are shown in succession.

At the same time the meter advances a test area to the tip of the cassette.



Do not touch the tip of the cassette while a test area is being advanced and do not approach the tip of the cassette with any objects.

Shortly after the test area has been advanced, the prompt *Apply drop* appears and a beep tone sounds again.

You can also open the tip cover if the meter is already turned on. In this case, the meter then switches to *Test*. There are the following exceptions:

- If you are in the *Settings* menu entering a number (for example, to change the time) or in a selection list (for example, to change the volume): In this case, when you open the tip cover you will be asked whether you really want to perform a test (see *Messages and problems* page 121).
- You had turned the meter on using the ▼ or ▲ button (see *Using the meter as a diary* page 65 or *Tools menu* page 103): In this case, the meter does not switch to *Test*. You must first turn the meter off. Then you can start the blood glucose test.
- You are in the *PC analysis* menu and have started an analysis from *Data transfer* or *Reports*.

B – Starting a blood glucose test from the Main menu:


1

If the meter is off:

Turn the meter on with .

Main menu appears on the screen.

2



Use ▼ ▲ to select *Test* .

Press .

If the tip cover is closed, the prompt *Open tip cover* appears on the screen.

3




Open the tip cover.


Press  if you want to cancel the test.



1



2



3

After the prompt *wash your hands* 1, a new test area is advanced 2 and the prompt *Apply drop* 3 appears.

Obtaining a blood drop

In general, you can obtain blood from any finger. Certain fingers can be unsuitable if, for example, a skin or fingernail infection is present. We recommend obtaining capillary blood from the sides of the fingertips as these areas are the least sensitive to pain.



- Use a new lancet each time you obtain blood. This reduces the risk of infection and blood collection remains virtually pain free.
- Only use the finger pricker when the cap is attached. If the cap is not attached, the lancet penetrates too deeply and the puncture may cause discomfort.



Press the finger pricker firmly against the selected puncture site.

Press the release button all the way down.

The lancet is primed and released in one step. The lancet penetrates the skin.

2

Massage the finger in the direction of the fingertip to encourage a blood drop to form.

Apply the blood drop to the test area immediately after you have obtained the blood drop.



Do not wait before applying the blood drop to the test area.

The amount of blood that emerges at the puncture site depends on the penetration depth and the pressure used to hold the finger pricker against the skin.


If not enough blood emerges, apply more pressure with the finger pricker the next time you obtain blood. If that is not sufficient, increase the penetration depth progressively too.

If too much blood emerges, reduce the penetration depth.

The blood drop must be large enough and well formed.

Applying the blood drop

You can hold the meter in your hand or lay it down while you apply blood to the test area.




Apply the blood drop to the centre of the test area.

Touch the blood drop to the test area gently.



- Do **not** hold the meter with the test area pointing upwards.
- Do **not** press the test area and the finger together.
- Avoid trembling or shaky movements.

The blood drop is drawn up by the test area.




Remove your finger from the test area when the beep tone sounds and *Test in progress* is displayed.

The test area has drawn up enough blood and the beep tone sounds. The hourglass symbol indicates that the test is in progress.



You can only hear the beep tones when they are turned on (see *Setting tones* page 49).



Testing is complete after approx. 5 seconds, depending on the blood glucose concentration. The test result appears on the screen and the beep tone sounds. The meter saves the test result at the same time. The used test area is transported away from the tip of the cassette.

If you want to flag the test result, do not turn the meter off (see *Flagging results* page 37).

Close the tip cover to turn the meter off.

Symbols associated with test results

The following symbols can be displayed together with the test result:

Symbol	Meaning
	The batteries are almost out of power.
	The temperature during the test was outside the permitted range of +10 to +40 °C.
	The test result is higher than the upper limit of the target range set.
	The test result is lower than the lower limit of the target range set.

Symbols instead of test results

The meter measures blood glucose values in the interval from 0.6 to 33.3 mmol/L. If the test result is outside this interval, one of the following symbols is displayed:


Symbol	Meaning
LO	The test result is lower than 0.6 mmol/L.
HI	The test result is higher than 33.3 mmol/L.



- The screen **LO** may indicate that your blood glucose value is very low (possibly a severe hypoglycaemia). The screen **HI** may indicate that your blood glucose value is very high (possibly a severe hyperglycaemia).
- Follow the relevant instructions given by your healthcare professional immediately and repeat the test.
- If you are using the acoustic mode: See page 91 for the beep tones for the symbols **LO** and **HI**.

Flagging results

You can flag results to describe certain events connected to this result or particular characteristics of the result. You can only flag a result while the result is still displayed after a test.



Press .

The *Set flag* menu is displayed. You can flag the test result here. There are 4 different flags to choose from. The symbol for the flag that has been activated is displayed to the right above the result. The test result in the above example is flagged with the symbol  *Other* **1**.

4

Performing a blood glucose test

You have the choice of the following 4 flags:

Symbol	Meaning
	<i>Before meal</i> (apple symbol): For test results that you obtained from tests before meals.
	<i>After meal</i> (apple core symbol): For test results that you obtained from tests after meals.
	<i>Other</i> (asterisk symbol): You can define the meaning of this flag yourself.
	<i>Control test</i> (applicator symbol): For control tests in which you applied control solution to the test area instead of blood.

You can add the following flags to a test result at the same time:


- *Other* and *Before meal*
- *Other* and *After meal*

You cannot flag a test result with *Before meal* and *After meal* at the same time.

You can change (add or remove) flags again as long as you are still in the *Set flag* menu. As soon as you leave the menu, the flags chosen are saved and can no longer be changed.

Selecting the *Control test* flag takes you out of the *Set flag* menu. The *Control solution* menu then opens (see Chapter 9, *Checking the meter*).

Selecting the *Before meal*, *After meal* or *Other* flag:



Press .



Use   to select *Before meal*, *After meal* or *Other*.


Press .

If you want to add a further flag to the test result, repeat step 2.

4


Performing a blood glucose test

Removing a flag again:



Use \blacktriangledown / \blacktriangle to select the flag, for example, *Other*.


The symbol displayed to the right above the test result shows which flag has been selected **1**.



Press \ominus to delete the flag.

The flag is deleted. The symbol on the screen disappears **2**.

Leaving the *Set flag* menu:



Turn the meter off with \ominus .

Or

Use \blacktriangledown / \blacktriangle to select *Main menu* and press \ominus .




Close the tip cover.

Setting a reminder

The *Reminder* option allows you to set a reminder from the *Set flag* menu straight after a test if you would like to check your blood glucose again later.


The meter can give you a one-time reminder to perform a test in 1 hour, in 1.5 hours, in 2 hours or in 3 hours. You do not need to set a time for this. Once the reminder has occurred, it is automatically deleted. For more information on reminders, see *Setting reminders* page 51.

When the test result is displayed:



Use \blacktriangledown \blacktriangle to select *Reminder*.

Press Ⓚ .



Use \blacktriangledown \blacktriangle to select a reminder time (for example, *In 1 hour*).

Press Ⓚ .

The meter returns to the *Set flag* menu.

Evaluating blood glucose results

Blood glucose results are influenced by, amongst other things, the type of food consumed, medication taken, state of health, stress and physical activity.



- Do not change your therapy without consulting your healthcare professional first.
- Consult your healthcare professional if the test result is below or above the blood glucose range you have set together with your healthcare professional.
- If the test result matches how you feel, follow the instructions given by your healthcare professional.
- If the test result does not match how you feel, for example, it is unexpectedly high or low, perform a control test with Accu-Chek Mobile control solutions. Then repeat the blood glucose test. If the new blood glucose result still does not match how you feel, contact your healthcare professional.
- Consult your healthcare professional immediately if your blood glucose values are too low or too high.
- The screen **LO** may indicate that your blood glucose is very low (possibly a severe hypoglycaemia). The screen **HI** may indicate that your blood glucose is very high (possibly a severe hyperglycaemia). Follow the instructions given by your healthcare professional immediately and repeat the blood glucose test.
- If test results repeatedly do not match how you feel, check the points listed in the following section *Implausible results - possible sources of error*.

Implausible results - possible sources of error

If your meter repeatedly displays implausible test results or error messages, check the points listed below. Answer the questions in the following overview with yes or no.

- If your answer to the question is **no**, make the respective corrections and repeat the blood glucose test.
- If you have answered **yes** to all these questions and still obtain implausible test results or error messages, contact your customer support and service centre (see page 134).

Source of error

Did you perform the blood glucose test according to the User's Manual?

Did you wash your hands with warm water and soap and dry them well?

Did you wait for the prompt on the screen before applying blood?

Did you apply the blood drop to the test area immediately?

Did you apply a well-formed blood drop?

Did you apply the blood drop to the centre of the test area?

In doing so, did you touch the blood drop to the test area gently?

Did you hold the finger and meter steady while applying the blood drop?

Did you remove the finger from the test area immediately after applying the blood drop?

Did you use a new blood drop for every test?

Did you perform the test within the correct temperature range (+10 to +40 °C)?

Did you observe the storage conditions for the meter and test cassette?

Did you pay attention to the sources of error stated in the test cassette package insert?

Did you perform a control test after your meter was dropped while *Apply drop* was being displayed?



If your meter has been dropped, this can also lead to implausible test results or error messages. In this case, you should also contact the customer support and service centre (see page 134).

Selecting settings

Settings overview

Menu	Available options
Language	Different languages
Volume	5 levels (<i>Mute</i> and 4 volume levels)
Tones	<i>Beep tones</i> : Turn the beep tones on or off <i>Reminders</i> : Turn the beep tone for reminders on or off <i>Acoustic mode</i> : Turn Acoustic mode on or off (see <i>Acoustic mode</i> page 89)
Reminders	7 times of day – freely selectable daily reminders
Time/Date	<i>Time</i> : Set the time <i>Date</i> : Set the date <i>Time format</i> : Set the time format (8 time formats) (format for time and date)
Target range	Set the upper and lower limit Turn the target range on/off
Brightness	3 levels



5

Selecting settings

The meter has several features which you can only use after making the appropriate settings. Most of these features are intended primarily to enable you to adapt the meter to your personal needs (for example, beep tones and acoustic reminders).

However, blood glucose tests can be performed independently of the settings. You can still perform a blood glucose test even if you have not made settings in the *Settings* menu.

To open the *Settings* menu:



In the *Main menu*, use \blacktriangledown \blacktriangle to select *Settings* (✓).

Press \odot .


The *Settings* menu is displayed.

From the *Settings* menu, you can open the following menus:

- *Language*
- *Volume*
- *Tones*
- *Reminders*
- *Time/Date*
- *Target range*
- *Brightness*

Setting the language


You can choose a language from the list for the text appearing on the screen.



In the *Settings* menu, use **▼▲** to select *Language* (🗣️).

Press **⏏**.


The language currently set is highlighted.



Use **▼▲** to select the desired language.

Press **⏏**.

The selected language is displayed.




Press **⏏**.

You have confirmed the language selection.

Correcting a wrongly set language

You can change the language if you selected the wrong one by mistake, as follows:




Turn the meter on with **⏏**.

Wait until the screen does not change anymore.

Press **▼** twice.

The *Settings* (🔍) menu is now highlighted.




Press **⏏**.

The *Language* (🗣️) menu is now highlighted.

Press **⏏**.

The *Language* menu is open.



Use **▼▲** to select the desired language.

Press **⏏**.

The selected language is displayed.


Press **⏏**.

The meter returns to the *Settings* menu.

Setting the volume

You can set the volume for the beep tones to one of 5 different levels.


The longer the bar on the screen, the louder the volume. At volume level 1 (*Mute*) no beep tone is audible.



In the *Settings* menu, use \blacktriangledown \blacktriangle to select *Volume*.

Press O .

The *Volume* menu is displayed.



Use \blacktriangledown \blacktriangle to set the volume.

Each time you press the button, a beep tone sounds in the new volume selected.

Press O .

You have confirmed the volume selection.

Setting tones

In the *Tones* menu you can turn the beep tones, the beep tone for reminders and the acoustic mode on or off. The different beep tone settings are as follows:

Beep tones: If *Beep tones* is turned on, a beep tone sounds during a test,

- when the meter advances a test area to the tip of the cassette,
- when the prompt *Apply drop* is displayed and you can apply blood or control solution to the test area,
- when the test begins and *Test in progress* is displayed,
- when the result is displayed,
- when an error message is displayed.

In order to be able to hear the beep tones, the volume must be set between volume levels 2 and 5. If you have set volume level 1 (*Mute*), no beep tones are audible. See *Setting the volume* page 48.


Reminders: If *Reminders* is turned on, a beep tone sounds for a set reminder (see *Setting reminders* page 51).

Acoustic mode: If acoustic mode is turned on, the meter guides you through the blood glucose test using the previously mentioned beep tones (see *Acoustic mode* page 89).

Turning tones on and off


You can turn the *Beep tones*, *Reminders* and *Acoustic mode* tones on and off.

Procedure using the example of *Beep tones*:



In the *Settings* menu, use ▼▲ to select *Tones*.

Press .



Use ▼▲ to select *Beep tones*.

Press  to turn the beep tones on () or off (.

When you have set the beep tones: Use ▼▲ to select *Back* or *Main menu*.

Press .

You have set the beep tones.



Note:

- If you want to use reminders, you must turn on both *Reminders* and *Beep tones*.
- If you want to use acoustic mode, you must turn on both *Acoust. mode* and *Beep tones*.


Setting reminders

You can set up to 7 reminder times.

If the meter is turned off, the *Reminder* message will be displayed daily at the time you have set. If you have turned on *Reminders* in the *Tones* menu, beep tones will also sound (see *Turning tones on and off* page 50).

Turning off the *Reminder* message beep tone

1



When the reminder is displayed, you can turn off the beep tones.


Press ▼ or ▲.

2

If you do not press a button, the meter turns off automatically after 20 seconds. The reminder is then only repeated on the following day.


If you perform a blood glucose test 10 minutes or less before a reminder, the reminder does not occur.

Setting a new reminder




In the *Settings* menu, use \blacktriangledown \blacktriangleleft to select *Reminders*.

Press O .



Use \blacktriangledown \blacktriangleleft to select *New*.

Press O .




Use \blacktriangledown \blacktriangleleft to set the hour **1**.

Press O .

Use \blacktriangledown \blacktriangleleft to set the minutes **2**.


Press O .



Only for times in 12-hour time format:


Use \blacktriangledown \blacktriangleleft to select *am* or *pm*.

Press O .



The meter shows you the set reminder time. The reminder is also turned on.

Press O .




The meter returns to the *Reminders* menu.

The set reminder time is displayed.

Turning a reminder on or off

A list of the reminders that have already been set appears on the screen.




In the *Settings* menu, use ▼▲ to select *Reminders*.

Press **⏏**.

You recognise an activated reminder by the ✓ symbol and a deactivated reminder by the ⋮ symbol.

Use ▼▲ to select the reminder time you wish to turn on or off.

Press **⏏**.




The menu option *On* is highlighted.

Press **⏏** to turn the reminder on (✓) or off (⋮).


A reminder that reminds you once to check your blood glucose value is not displayed in the list (see *Setting a reminder* page 41).

Changing a reminder




In the *Settings* menu, use ▼▲ to select *Reminders*.

Press .



Use ▼▲ to select the reminder time you wish to change.

Press .




Use ▼▲ to select *Edit*.

Press .

Set the new reminder time.


The reminder time has now been changed.

Deleting a reminder




In the *Settings* menu, use ▼▲ to select *Reminders*.

Press .



Use ▼▲ to select the reminder time you wish to delete.

Press .



Use ▼▲ to select *Delete*.

Press .


The reminder time has now been deleted.

Setting the time, date and time format

The time and date are preset in the meter. If you want to use reminders or analyse your test results with a computer, for example, you should check the time and date and correct them if they deviate from the current time or date.

All test results are saved together with the time and date. It is only possible to analyse the test results meaningfully in terms of time if the time and date are set correctly.

Setting the time




In the *Settings* menu, use ▼▲ to select *Time/Date*.


Press .

Use ▼▲ to select *Time*.


Press .



Use ▼▲ to set the hour ①.
Press .



Use ▼▲ to set the minutes ②.
Press .



Only for times in 12-hour time format:

Use ▼▲ to select *am* or *pm*.

Press .



The set time is displayed.

Press .


You have confirmed the time set.

Setting the date

The sequence in which you set the day, month and year while setting the date depends on the time format you have set. Apart from the sequence, the procedure is the same in all cases.

The following sequences are possible:

Day.Month.Year	Month/Day/Year	Year/Month/Day
		




In the *Settings* menu, use ▼▲ to select *Time/Date*.

Press **⊙**.


Use ▼▲ to select *Date*.

Press **⊙**.




Use ▼▲ to set the day **1**.

Press **⊙**.




Use ▼▲ to set the month **2**.

Press **⊙**.



Use ▼▲ to set the year **3**.


Press **⊙**.




The set date is displayed.

Press **⊙**.

You have confirmed the date set.



Setting the time format




In the *Settings* menu, use ▼▲ to select *Time/Date*.

Press **Enter**.

Use ▼▲ to select *Time format*.


Press **Enter**.

The currently set time format is highlighted.



Use ▼▲ to select the desired time format.

Press **Enter**.



The selected time format is displayed.

Press **Enter**.

You have confirmed the selected time format.

If you do not want to set the selected time format, use ▼▲ to select *Cancel*. Press **Enter** to return to the *Time/Date* menu.

Setting the target range

You can enter your personal target range for blood glucose values (default setting: Off). Test results above this target range are flagged with the symbol \uparrow . Test results below this target range are flagged with the symbol \downarrow .


The target range comprises the blood glucose values which should be achieved if the therapy is optimal. Consult your healthcare professional for the target range appropriate for you.

You set the target range by entering the lower and upper limits. The lower limit can be set to between 2.8 and 5.5 mmol/L. The upper limit can be set to between 5.5 and 11.1 mmol/L. Both limits can be set in steps of 0.1 mmol/L.




If you want to use the lower limit of the target range as an indicator of a possible hypoglycaemia (low blood glucose), it is important to remember the following: The indicator is only reliable if the limit has been properly selected. We therefore strongly recommend that you only set the limit in consultation with your healthcare professional. This function is not a substitute for hypoglycaemia training by your healthcare professional or diabetes team.

Setting a target range




In the *Settings* menu, use \blacktriangledown \blacktriangle to select *Target range*.

Press \odot .



Use \blacktriangledown \blacktriangle to select *Edit*.

Press \odot .



Use \blacktriangledown \blacktriangle to set the lower limit

1.


Press \blacksquare .

Use \blacktriangledown \blacktriangle to set the upper limit

2.

Press \blacksquare .


Turning the existing target range on/off



Press \blacksquare .

The set target range is displayed.
The target range is also turned on.


If you do not want to set a target range, use \blacktriangledown \blacktriangle to select *Cancel* and then press \blacksquare to return to the *Settings* menu.



Use \blacktriangledown \blacktriangle to select *Back*.


Press \blacksquare .

The meter returns to the *Settings* menu.



In the *Settings* menu, use \blacktriangledown \blacktriangle to select *Target range*.

Press \blacksquare .




Use \blacktriangledown \blacktriangle to select *On*.

Press \blacksquare to turn the target range on (\checkmark) or off (\dots).


You recognise an activated target range by the \checkmark symbol and a deactivated target range by the \dots symbol.

Changing a target range




In the *Settings* menu, use ▼▲ to select *Target range*.

Press **OK**.



Use ▼▲ to select *Edit*.

Press **OK**.



Set the new target range.


The changed target range is displayed.

Press **OK** to confirm the set target range.

The target range is also turned on.


Setting the brightness

You can set the brightness of the texts and symbols appearing on the screen to one of 3 different levels. The more bars filled on the screen, the brighter the screen.



In the *Settings* menu, use ▼▲ to select *Brightness*.

Press .



Use ▼▲ to set the desired brightness.

Press .

You have confirmed the selected brightness.

Using the meter as a diary

The meter can store up to 2,000 results with the time and date. The meter stores the following information:

- all results,
- all symbols displayed with the result,
- all flags that were set.


If all memory locations are occupied, the oldest result is deleted when you perform a new test, to create space for the new one.

The following options are available in the *My data* menu:

- *All results*: Display all the stored results
- *Averages*: Display the averages for the last 7, 14, 30 or 90 days
- *PC analysis*: Display and analyse the stored test results on the computer (PC) (see *Analysing test results on the PC* page 70)

Retrieving stored results


There are 2 ways to retrieve stored results: By opening the *My data* menu (A) or by pressing the ▼ button (B).

A – Retrieving stored results via the *My data* menu:


In the *Main menu*, use \blacktriangledown \blacktriangle to select *My data* (☞).

Press \blacksquare .

The *My data* menu is displayed.



Use \blacktriangledown \blacktriangle to select *All results*.
Press \blacksquare .



The last saved result is displayed.

Symbols saved together with the result appear to the right above the result.

B – Retrieving stored results using the \blacktriangledown button:

1

With the meter off: Press and hold \blacktriangledown until the meter turns on (longer than 2 seconds).

The last saved result is displayed.

C – Retrieving older results:

1

Retrieve the stored results (see page 65).

2

Press ▼ to go to the next oldest result.

Press ▲ to go to the next newest result.

If you press and hold the ▼ or ▲ button, the results are displayed in quick succession, starting with a slow speed and becoming faster. The screen stops at the oldest and the newest (latest) result.

No results stored





If the last (most recent) result in the memory is displayed as *XX.X mmol/L* with dashes instead of the time and date, the last test did not deliver a result. This means that during the last test either an error message was displayed or you did not apply blood (or control solution) after the prompt *Apply drop* was displayed. After the next test that delivers a result, the result displayed with X is deleted from the memory.

XX.X mmol/L is also displayed if no results are stored in the meter.

Retrieving averages

The meter can calculate averages from the stored test results.



You can choose which test results are used for calculating the averages:

- All results **1**
- Only test results that were flagged with *Before meal* **2**
- Only test results that were flagged with *After meal* **3**

When you have made this selection, you can choose the period of time in which the test results were obtained. The last 7, 14, 30 or 90 days.

The average is calculated from all test results that match your selection. However, the following results are not included in the calculation:

- Control tests (flagged with )
- Test results displayed as **LO** or **HI**

Selecting the time period **4** only determines which average is displayed first. You can use the   buttons to switch directly from one time period to the others. The calculated average **5** is displayed, together with the number of test results (called Tests) from which the average was calculated **6**.

1



In the Main menu, use ▼▲ to select *My data* (⊞).

The *My data* menu is displayed.

2



Use ▼▲ to select *Averages*.

Press ⊞.


3



Use ▼▲ to select *All results*, *Before meal* or *After meal*.


Press ⊞.

4



Use ▼▲ to select the time period in which the test results were obtained: *7*, *14*, *30* or *90 days*.

Press ⊞.



Averages *Before meal* are marked with the symbol  1. Averages *After meal* are marked with the symbol  2.



- The meter calculates the average for a time period even when the stored test results cover a shorter period (for example, only 5 days).
- If no averages can be calculated, *XX.X mmol/L* is displayed instead of a value and the number of test results (tests) is given as zero. This is the case if either no test results or only test results which are not used for calculation of averages were saved in the specified time period. The following test results are not taken into account for the calculation of averages:
 - Test results outside the time period selected
 - Test results without date and time
 - Tests with control solution
 - Tests outside the measuring range (marked with *HI* or *LO*)

Analysing test results on the PC

The meter has a built-in USB port (Universal Serial Bus) ① for transferring stored test results to an appropriately equipped computer (PC).

USB port



You need 1 USB cable to connect the meter to a PC. The USB cable must have the following connectors:

- Micro B connector ② to establish a connection with the meter.
- USB A connector ③ to establish a connection with the PC.


There are **two ways** of displaying the test results on a PC via the *PC analysis* menu:

- **Reports** allows you to display and analyse the test results on a PC by means of the meter's built-in diary function software and an Internet browser, no additional software is required (see *Reports* page 77).
- **Data transfer** transfers the test results to the PC. You can use special software products for diabetes management (for example, Accu-Chek 360° Diabetes Management System or Accu-Chek Smart Pix software) to display the data on a PC and analyse it (see *Data transfer* page 74). With this setting, the meter also supports data exchange between medical devices with the Continua Certified® logo from the Continua Health Alliance.

These analysis options help you and your healthcare professional to manage your data and the graph and table views help you to better understand your test results.


Defining the default

You can define how the meter reacts after being turned on when you connect it to a PC.




In the Main menu, use ▼ ▲ to select *My data*.

Press .



Use ▼ ▲ to select *PC analysis*.

Press .



Use ▼ ▲ to select *Default*.

Press .

4



Use **▼ ▲** to select *Data transfer*

or

Reports.

Press **Ⓞ**.

Depending on the default you select, the analysis of the test results is started with *Data transfer* or *Reports*. To analyse the data by means of *Data transfer*, you need a special software product for your PC (for example, Accu-Chek 360° Diabetes Management System or Accu-Chek Smart Pix software). For more information, contact the customer support and service centre (see page 134).

If you already have a software product for transferring and analysing test results, the software may not recognise more recent meters and therefore the test results may not be able to be transferred. You may need a more recent version of your software. In this case, contact the customer support and service centre (see page 134).

You cannot perform a test while test results are being transferred.



You only need special diabetes management software to analyse stored test results if you would like to use the data transfer function. For transferring data between Continua Certified® products, you may also need to install a PHDC driver (PHDC = Personal Health Device Class) on your PC. Whether you need a PHDC driver depends on the diabetes management software used.

Viewing a report (Accu-Chek Mobile report) with your Internet browser does not require any additional software.

Connecting the meter to the PC

1



Plug the Micro B connector of the USB cable into the USB port on your meter.

Connect the other end of the USB cable to your PC.

2a and 2b


2a – With the meter off:

The meter turns on. Depending on the selected default (see *Defining the default* page 71), the analysis of the test results is started with *Data transfer* or *Reports*.

2b – With the meter on:

The meter opens the *PC analysis* menu. Other running operations are aborted.

2c



If you connect the meter to a PC while a test is in progress, the test is cancelled and a message appears on the screen.

Press .


The meter starts data analysis via *Data transfer* or *Reports*, depending on the default.



- If you disconnect the cable between meter and PC, the meter turns off.
- Remove the USB cable after analysing the test results on the PC so that the meter battery power is not consumed unnecessarily.
- You cannot perform a test while the meter is connected to the PC.

Starting data transfer

1



2

If necessary, start the software program for test result analysis on the PC.

When transfer of the test results is finished, the meter turns off automatically.


Make sure the meter is connected to the PC (see [Connecting the meter to the PC](#) page 73).

While the connection is being established and data is being transferred, you will be informed of the progress of these operations on the screen.



- To analyse the data by means of *Data transfer*, you need additional software (for example, Accu-Chek 360° Diabetes Management System or Accu-Chek Smart Pix software).
- If you do not want to transfer the test results after all, you can abort the transfer using the  button. The meter returns to the *My data* menu (if you came from the *My data* menu) or turns off (if you turned the meter on using the  and  buttons).

Displaying reports



Make sure the meter is connected to the PC (see [Connecting the meter to the PC page 73](#)).

While the connection is being established, you are informed of the progress on the screen. The PC symbol on the meter screen flashes when the meter has successfully connected to the PC.

2

Open the file manager of your operating system.

The meter appears as a drive (USB storage device) in the file manager.

3


Double click on the (ACCU-CHEK) drive icon to open it.

Double click on the *Start.html* file to open it.

The Internet browser opens and the default reports appear.

Analysis with a meter already connected


If you have already connected the meter to a PC with the USB cable but the meter is off, proceed as follows:



Turn the meter on with .


In the Main menu, use   to select *My data*.

Press .



Use   to select *PC analysis*.

Press .



Use   to select *Data transfer* or *Reports*.

Test results start transferring when you select *Data transfer*.

The meter connects to the PC when you select *Reports*.

Press  to confirm your selection.

Reports

Reports are used to easily and automatically analyse stored blood glucose results from your Accu-Chek Mobile meter.

Overview

The reports can be displayed on a PC using an Internet browser and can be printed by a printer. You do not need an Internet connection to display the reports. The reports are stored in the meter and retrieved from there.



You can find the license agreement for using the meter's built-in *Reports* software at the end of the User's Manual (see page 142).

When you have successfully connected the meter to the PC, the following reports are displayed on a page of the Internet browser:

- *Trend Report*
This report shows the trend of several blood glucose results over the selected time period.
- *Standard Day Report*
This report displays all data in a 24-hour grid.
- *Standard Week Report*
This report displays all blood glucose values according to the time when the test was performed and the respective day of the week.
- *List Report*
The List Report shows the test results sorted by date and time of the test.

User interface

Accu-Chek Mobile report

Trend Report 27.04.2014 – 10.05.2014

Blood Glucose (mmol/L)

27 April 2014 May 11

Statistics

Number of levels:	90	Average tests per day:	6.4	Above target range	>7.8 mmol/L	46.7% (42)
Average (mmol/L)	8.03	HI:	0	Within target range	3.9–7.8 mmol/L	27.8% (25)
Standard deviation (mmol/L)	4.88	LO:	0	Below target range	0.0–3.9 mmol/L	25.6% (23)
Highest value (mmol/L)	22.2	Low blood glucose index:	7.0			
Lowest value (mmol/L)	1.7	High blood glucose index:	8.2			

Key

x Measurement	^ Result(s) above graph	▨ Blood glucose target range
x Multiple measurements	^ HI	▨ Non-work day
— Average	∨ LO	

Control elements:

- 1 Window of the Internet browser
- 2 Drop-down menu for selecting the time period
- 3 *Print reports* button
- 4 *Save file* button

Display panes:

- 5 Legend for explanation of symbols
- 6 Statistics on test results analysis
- 7 Chart (graphic representation of the report)
- 8 Time period



The test results are represented in the unit **mg/dL** or **mmol/L**, depending on the version of the meter.

Symbols

Within the report charts, test results are plotted in the analysis by means of different symbols.


The symbols and graphic elements in the charts have the following meaning:

Symbol	Meaning
	<i>Test:</i> Blood glucose value from a test
	<i>Several tests:</i> Blood glucose values from several tests
	<i>Average:</i> Average of the blood glucose values in the selected time period
	<i>Target range for blood glucose values:</i> Personal target range for blood glucose values (represented on the screen as a green bar)
	<i>Test result(s) above chart:</i> The blood glucose value cannot be represented because it is above the chart
	<i>Non-work day:</i> Usually non-working days (weekends)

Report time periods

Reports can be created for the following time periods:

- *Past 3 days*
- *Past 7 days*
- *Past 14 days*
- *Past 30 days*
- *Past 90 days*



Select the desired time period in the drop-down menu (for example, *Past 14 days*).

Printing reports

All the reports displayed on the screen can also be printed out together.

Do not use the integrated print function of the browser to print the reports, instead use the *Print reports* () button on the user interface.

1

Click  on the user interface.

The print dialogue box of your Internet browser opens.

2

Select the desired printer from the print dialogue box of the Internet browser.

In the print dialogue box, you have the option of making further settings for the print process.

3

Select whether the report is to be printed in portrait or landscape format.

Start the printing process.

Depending on the operating system and system configuration, the procedure for selecting the print settings on your PC may vary.

Analysed data

The following data is not included in the statistical analysis:


- Test results outside the selected time period
- Test results without date and time
- Tests with control solution
- Tests outside the measuring range (marked with *HI* or *LO*)

The following settings made in the blood glucose meter are adopted for data analysis:

- Language (see page 47)
- Date (see page 57)
- Time format (see page 59)

Standard Day Report


Standard Day Report



The *Standard Day Report* makes it easier to recognise daily patterns. For this purpose, all the data is placed on a 24-hour grid. As a result, all tests performed at (approximately) the same time of day are shown at the same position on the time axis.

All blood glucose results are plotted at the relevant time of day according to the time when the test was performed. The values are connected by a thin black line in chronological order. A thick black line represents the trend of the average level (in intervals of one hour if a test result falls in each interval).

Standard Week Report



The *Standard Week Report* makes it easier to recognise repeated patterns depending on the day of the week. For example, it may be possible to recognise changes in metabolic situation brought about by your lifestyle patterns.

All blood glucose results are plotted on the chart according to the time when the test was performed and the respective day of the week. The values are connected by a thin black line in chronological order. A thick black line represents the trend of the average level on the different days of the week.

The number of tests and the average for the respective day of the week are listed below the chart.

List Report

List Report

27.04.2014 – 10.05.2014

Date and Time	Blood Glucose (mmol/L)	Events
09.05.2014 07:30	11.7	Before meal
01:01	9.9	
08.05.2014 20:00	9.4	After meal
17:45	3.9	Before meal
15:00	3.6	After meal
12:27	2.2	Before meal
10:18	16.6	After meal
07:37	17.2	Before meal
01:10	10.4	
07.05.2014 22:00	3.9	After meal
20:10	3.9	After meal
17:27	4.7	Before meal
15:17	3.7	After meal
12:25	1.8	Before meal
10:05	16.6	After meal
07:41	17.6	Before meal

The *List Report* shows the test results sorted by date and time of the test. All blood glucose values are listed chronologically and with any additional information about events (see *Flagging results* page 37).

The list contains the following columns:

- Date and Time
- Blood Glucose (mmol/L)
- Events
 - Event connected to this test result

Statistics

Below the chart of a report, you will find a statistical analysis of all plotted test results with the following information:

- Number of tests
- Average (mmol/L)
- Standard deviation (mmol/L)
The standard deviation is the variance of the analysed results.
- Highest blood glucose value (mmol/L)
- Lowest blood glucose value (mmol/L)
- Average tests per day
Average number of blood glucose results per day
- HI or LO
Blood glucose results outside the measuring range
- Low blood glucose index or High blood glucose index
Further information can be found in the appendix (see page 140).
- Above target range
Blood glucose results above the target range
- Within target range
Blood glucose results within the target range
- Below target range
Blood glucose results below the target range

Analysing data in external applications

If you want to analyse the test results using external software, you can save the data as a CSV file (CSV = Comma Separated Values). CSV files can be opened with a text editor or spreadsheet program, for example.

The CSV file always contains all test results saved in the meter. Test results transferred at an earlier date are transferred again.

1

Click  on the user interface.

Depending on the configuration of your PC, the CSV file may be directly opened in a spreadsheet program. In this case, you can save the table using the Save function of the spreadsheet program.

2

In the dialogue box that opens, select the option to save the file.

Where you find the CSV file on your computer depends on the settings in your operating system for downloading files.

The CSV file contains the following information:

- *Serial number* of the meter.
- *Download date, download time* – date and time when the meter transferred the test results to the computer.
- *Date, time, result* and *unit* of the results saved in the meter.
- *Flags* added to the results, indicated by an *X*.



In the CSV file, the date is always displayed as DD.MM.YYYY and the time as 24-hour format (hh:mm). The time format set in the meter has no influence on the format of the date and time in the CSV file.

Security settings in the Internet browser

The settings of the Internet browser can influence working with reports.

The reports use pages with so-called *active content* (JavaScript). This *active content* can be suppressed by security settings in the browser. If you suppress JavaScript, it may cause warnings or restricted functionality. For this reason, check the relevant JavaScript settings in the Internet browser to ensure smooth operation.

In many cases, you can create different security settings for using the Internet and for working with reports (for example, at user login to the PC or by defining user profiles in the browser).

If you select the Internet browser security settings appropriately (for example, *Allow active content to run in files on My Computer* or similar), you can work with reports without any restrictions.

Error messages and troubleshooting

Error message	Troubleshooting
The ACCU-CHEK drive symbol with the <i>Start.html</i> file does not appear on the PC:	Check whether <i>Reports</i> is selected in the defaults for PC analysis (see <i>Defining the default</i> page 71).
	Check whether your PC or operating system supports data transfer via USB.
	Check whether the USB connector is firmly plugged into the correct socket on the PC.
If the meter is still not detected as a drive:	Plug the meter into a different USB socket on your PC (you may be able to use a USB hub or choose between USB sockets on the front and rear of your PC).

Contact the customer support and service centre (see page 134) if the problem persists.

Acoustic mode

The procedure for turning the acoustic mode on is described in *Setting tones* (see page 49).

When the acoustic mode is turned on, the meter guides you through the blood glucose test using beep tones and announces the test result as a series of beep tones. The meter also beeps when the batteries are almost empty, the test cassette is empty or an error message is displayed.

An intimate knowledge of how test results are announced with beep tones and how to distinguish beep tones for test results from other beep tones, is crucial to interpreting the test result correctly.

The sections that follow explain the beep tones in detail. In addition, training software (the BeepLearn program) to help you interpret test results announced in the form of beep tones is available on CD from the customer support and service centre (see page 134).



Visually impaired people should only use the acoustic mode if they have familiarised themselves fully with the acoustic mode with the assistance of a sighted person and have demonstrated that they are able to understand test results correctly without exception, using the acoustic mode.

The acoustic mode should always be used with great care. Otherwise, there is a risk that the test results will be misinterpreted by the person performing the test. An incorrect insulin dose resulting from a misinterpreted test result can cause considerable damage to health and can even be fatal. If unexpected and abnormal blood glucose values are obtained using the acoustic mode, the test result must be checked by repeating the test together with a sighted person.

Visually impaired people must be assisted by a sighted person when performing a blood glucose test.

The meter must be checked regularly by a sighted person for outwardly visible contamination.

There are two types of beep tones, differing in pitch. Warnings have a higher pitch (Tone 2, represented herein as **—**) than the other beep tones (Tone 1, represented herein as **—**). The following table shows which tone sounds for which event.

Beep tones	
Tone 1: —	Tone 2: — (warning)
Turning on	Announcement of an error message
Test area is advanced to the tip of the cassette	Notification that the batteries are almost empty
Prompt to apply blood or control solution	Notification that the test cassette is empty
Start of test	
Signal that test results are about to be announced	
Announcement of the test result	

Beep tone when turning on

When the acoustic mode is turned on, you hear 1 beep tone (**—**) when you turn the meter on, regardless of how you turn the meter on.

Beep tones during a test

After the power-on beep, you hear 1 beep tone (**—**) during the test when:

- the meter advances a test area to the tip of the cassette,
- you are prompted to apply blood or control solution,
- the test begins.

You hear the test result, which is composed of different beep tones.

You hear several beep tones when the meter displays an error message (**—**).

Test result announcement after a test

In the acoustic mode, when a test result is displayed following a blood glucose test, the meter also announces it as a series of beeps. The test result is not announced as an entity, but is broken down into individual digits.

Examples:

The test result 7.6 mmol/L is announced as 0 – 7 – decimal point – 6.

Each digit is represented by the corresponding number of beep tones, for example, 4 beep tones for the number 4 (— — — —). Zero is represented by 1 long beep tone (————).

The decimal point of a test result is announced by 1 very short beep tone (≡).

Test results which are lower than 0.6 mmol/L and are displayed as **LO** are announced acoustically as 0 – 0 – decimal point – 0.

Test results which are higher than 33.3 mmol/L and are displayed as **HI** are announced acoustically as 9 – 9 – decimal point – 9.

The meaning of the symbols **LO** and **HI** is explained in the chapter *Symbols instead of test results* on page 36.

The test result is announced three times in succession. Each announcement is preceded by 2 short beep tones (≡≡) as an introduction. Therefore, altogether you hear: ≡≡ test result — test result ≡≡ test result.

Symbols, such as \uparrow or \downarrow , which are displayed together with the test result are not announced acoustically.

If you turn the meter off using the power button (⏻) while the test result is being announced acoustically, the announcement is interrupted and 1 long beep tone will sound (————).

mmol/L meters

First the tens are announced, then the units followed by the decimal point and finally the tenths. There is a short pause between each group of beeps. The tens are always announced, even when the test result is below 10. The tens in this case are represented by 1 long beep tone, signifying zero. The decimal point is announced by 1 very short beep tone (≡).

13.8 mmol/L:

2 short preceding beep tones – 1 beep tone – pause – 3 beep tones – pause – 1 very short beep tone – pause – 8 beep tones = 

4.0 mmol/L:

2 short preceding beep tones – 1 long beep tone (for 0 tens) – pause – 4 beep tones – pause – 1 very short beep tone – pause – 1 long beep tone = 

Announcing stored test results

Only the last (most recent) of the stored test results is announced acoustically.

The test result is announced three times in succession. The announcement follows the same pattern as the announcement of a test result immediately after the test.

You will hear:  test result  test result  test result.

If you directly retrieve stored test results using the button , you will first hear the power-on beep and immediately afterwards the test result:

— — — test result — — test result — — test result.

If XX.X is displayed as the last (most recent) test result, the last test did not deliver a test result (see page 66). In this case, the beep tones issued are identical to those for an error message.

Flags are not announced acoustically (for example,  *Before meal*).

There is no acoustic announcement if the last saved result is a control test flagged with , if there are no stored results or if you retrieve an average.

Announcing warnings and error messages

When battery power is low or the test cassette has been used up, the meter issues an acoustic warning. This consists of 2 short beep tones () sounded three times in succession ( —  — ).

The warning is given at different times.

- Batteries almost empty:
The acoustic warning sounds after you have turned the meter on, together with the corresponding message on the screen (see *Changing the batteries* page 114).
- Test cassette empty:
The acoustic warning sounds when the number of available tests is displayed as zero when turning the meter off.

If error messages appear on the screen (see *Error messages* page 124), these are announced by 2 short beep tones () sounded four times in succession ( —  —  — ). This sequence is sounded only once and is not repeated. The same sequence of beep tones is sounded for all error messages, i.e. they are not differentiated acoustically.

Checking the meter

You can check whether the meter is delivering correct test results. To perform this control test, a glucose control solution is applied to the test area instead of blood. At the end of the test, the meter automatically checks whether the control result obtained with control solution is correct and informs you of the control result.

Always perform a control test using Accu-Chek Mobile control solutions:

- After you have cleaned the inside of the tip of the cassette and tip cover, as well as the measuring optics and the front opening of the meter.
- If you are in doubt about a test result.

Contact the customer support and service centre (see page 134) to find out where you can obtain the control solutions.



Visually impaired people must be assisted by a sighted person when performing a control test.

Different control solutions are sold in different countries. The control solutions have labels with different text colours (Control1 = blue 1 or Control2 = red 2).

At the end of the control test, you must tell the meter which text colour of the control solution you have used (see page 97). The meter offers you a choice of two colours.

Preparing for a control test

For a control test, you need:

- The meter with inserted test cassette
- An unopened applicator with Accu-Chek Mobile control solution
- A clean, dry paper towel.

The solution in the applicators is intended for single use only.

[Read the control solution package insert.](#)

Performing a control test


A control test in its main steps is the same as a normal test except that you apply control solution to the test area instead of blood.

1

Either open the tip cover or select *Test* in the Main menu.


Shortly after the test area has been advanced to the tip of the cassette, the prompt *Apply drop* appears and a beep tone sounds.

2



Twist the cap off the applicator.

3



Hold the applicator at a slight angle with the brush pointing downwards.


Squeeze the applicator gently until you see a small drop appear on the brush.

4



Apply the control solution to the centre of the test area at the tip of the cassette using the tip of the brush.


5



Take the brush away from the test area as soon as *Test in progress* appears on the screen.

The test area has absorbed sufficient control solution, the beep tone sounds and the control result is displayed.

6




Press .


The test ends after 5 seconds and the meter displays the result.




Do **not** hold the meter with the test area pointing upwards while you apply control solution because the solution could enter the meter and soil it.




Use **▼▲** to select *Control test*.
Press **⏏**.



Use **▼▲** to select the colour of
the text on the applicator label
(see page 94).
Press **⏏**.




The control result flagged as  for Control test is displayed.
Press **⏏**.



The meter displays the upper **1** and lower **2** limits of the concentration range. The concentration range displayed depends on the colour of control solution selected.

Note: The values in the illustrations are **only** examples.

Press .



The meter automatically checks whether the control result is within the concentration range displayed.

If the control result is within the concentration range displayed, *Control test OK* is displayed.

If the control result is not within the concentration range displayed, *Control test not OK Repeat control test* is displayed (see *Possible sources of error* page 99).


Press .

The meter returns to the Main menu.

Turn the meter off and close the tip cover.

Discard the used applicator according to local regulations.

Checking the concentration range yourself



You can also check yourself whether the control result is within the right concentration range.

To do this, compare the control result with the concentration table on the test cassette box.

The control result must be within the printed concentration range ②.

Note: The values in the illustrations are **only** examples.

① = unit of measurement, ② = concentration range, ③ = control solution used

Possible sources of error

If *Control test not OK Repeat control test* is displayed, the control result is not within the concentration range displayed. Answer the questions in the following overview with yes or no.

- If your answer to the question is **no**, make the respective corrections and repeat the test.
- If you answered **yes** to all the questions and *Control test not OK* is still displayed, contact the customer support and service centre (see page 134).



If *Control test not OK Repeat control test* is displayed, you can no longer guarantee that the meter and test cassette are functioning properly. Blood glucose tests may then produce incorrect test results. Incorrect test results can cause the wrong therapy recommendation to be made and thus produce serious adverse health effects.

Source of error

Did you perform the control test according to the User's Manual?

Did you wait for the application prompt to be displayed before applying control solution?

Did you apply the control solution immediately after opening the applicator?

Did you enter the correct colour in the *Control test* menu?

Did you test within the correct temperature range (+10 to +40 °C)?

Did you check the use by date for the control solution? Is the control solution still valid (use by date: see label on the bottom of the applicator; next to the  symbol)?

Did you observe the section *Storing and using control solutions properly* in the package insert for the control solution?

Did you observe the storage conditions for the meter, test cassette and control solutions (see *Measurement and storage conditions* page 117, and the test cassette and control solution package inserts)?

Did you use the meter properly? The meter has **not** been dropped? If your meter has been dropped, this can also lead to control results which are not within the concentration range displayed and for which *Control test not OK* is displayed. In this case, contact the customer support and service centre (see page 134).

Tools menu


On the one hand, the *Tools* menu offers you some extra means of checking the meter. On the other hand, you must also use this menu to prepare for cleaning the meter and changing a test cassette that has not been used up.

The following options are available in the *Tools* menu:

- *Display check*: Performs a separate display check
- *Validity*: Displays the validity of the test cassette
- *Change cass.:* Prepares the meter for removing a test cassette which still has unused test areas (for example, when you clean the meter).


Performing a display check

If you want to check whether the display is working properly, you can do so via the display check in the *Tools* menu, in addition to the display check after turning the meter on. In this case, the display check is displayed not only for about 2 seconds, but until you cancel it (1 minute at most).




In the Main menu, use \blacktriangledown \blacktriangle to select *Tools* (\blacklozenge).

Press \blacksquare .




Use \blacktriangledown \blacktriangle to select *Display check*.

The *Display check* is displayed.





Press \blacksquare .

The meter runs the display check in which a black and yellow checkerboard pattern of small illuminated dots is displayed.




Check whether there are any irregularities in the checkerboard pattern.




Displaying the validity of the test cassette

You can check how long the test cassette in the meter is valid for as follows:




In the Main menu, use   to select *Tools* ().

Press .



Use   to select *Validity*.

The validity is displayed.



Press .

The meter returns to the *Tools* menu.

Viewing the number of available tests

With the meter off:

1

Press and hold  until the meter turns on (longer than 2 seconds).

The number of available tests is displayed. The meter turns off again a short time later.

Replacing the test cassette

If you want to replace an empty test cassette or a test cassette that has expired, continue reading in the section *Replacing the test cassette* (see page 105).


If you want to remove a partly used test cassette which still has unused test areas, you must first perform the preparatory steps described in the section *Preparing test cassette replacement* (see page 104).



If you re-insert the partly used test cassette in the meter and there was another cassette in the meter before this, the partly used test cassette loses a test area.


Preparing test cassette replacement

The preparatory steps for test cassette replacement are only necessary if you want to remove a partly used test cassette which still has unused test areas:



In the *Tools* menu, use ▼▲ to select *Change cass.*.

Press .



The meter now prompts you to open the tip cover. If you do not want to replace the test cassette after all, you can cancel the operation with .

You can re-insert a partly used test cassette in the meter at a later date. In the meantime, store this test cassette in a dry place away from light.


Replacing the test cassette

Test cassette empty/defective/expired



If the test cassette in the meter is empty **1** or defective **2** or has expired **3**, the appropriate message appears after the display check and the number of available tests (see also page 120). Shortly after this, the meter turns off automatically.


Change cassette



If you have prepared for the test cassette replacement via *Change cass.* the meter prompts you to replace the test cassette **4**. Shortly after this, the meter turns itself off automatically.



Open the tip cover.



Turn the meter over so that the back is facing upwards.



Push the slide button for the cassette compartment cover upwards.

The cassette compartment cover opens.

4




Remove the old test cassette from the cassette compartment.

5



Take the new test cassette out of the plastic container.

6



Insert the new test cassette in the meter with the silver side facing upwards.

7



Close the cassette compartment cover.

8



Press the cassette compartment cover closed.

The cassette compartment cover must close with an audible **CLICK**.

The meter turns on and runs a display check. Afterwards the validity of the test cassette and the number of available tests are displayed.

Then the main menu is displayed.



Used test cassettes can be disposed of in household waste if no other regulations apply locally.

Cleaning the blood glucose monitoring system

Cleaning the meter

The mode of operation of the meter with the test cassette means that the meter will not normally come into contact with blood or control solution. Regular cleaning is therefore not necessary. In the event of the meter becoming soiled through improper use, cleaning may become necessary.



Visually impaired people must be assisted by a sighted person when cleaning the meter.



- Use **only** cold water or 70 % isopropanol for cleaning. Any other cleaning agents may damage the meter or impair its measuring function.
- Use a slightly dampened cloth or a slightly dampened cotton swab. Do not spray anything directly onto the meter and do not immerse it in cleaning fluid.

The outside

If the casing of the meter or the display is visibly soiled:

1

Wipe the casing or display with a cloth slightly dampened with cold water.


2

Repeat step 1 with a cloth slightly dampened with 70 % isopropanol.

The inside


You only need to clean the inside of the meter or the inside of the tip of the cassette when the message with the text *Cassette dirty: Clean per user's manual (E-4)* is displayed. The error message may be caused by dirt inside the tip of the cassette.

If you are prompted to clean the meter by the message *Cassette dirty: Clean per user's manual (E-4)*, proceed as follows:



Press .

Open the tip cover.




Follow the instructions (steps 3 to 11) for cleaning the meter in the section *Cleaning* (see page 109).



If the error message (E-4) was displayed, a test area is lost when you re-insert the test cassette later on.

Cleaning


If you notice that the area around the tip of the cassette is visibly soiled and you want to clean the meter, proceed as follows:



In the *Tools* menu, use \blacktriangledown \blacktriangle to select *Change cass.*.

Press \blacksquare .


The meter prompts you to open the tip cover.



If you do not want to clean the meter now, press \blacksquare .

If you do want to clean the meter, open the tip cover.

The meter prompts you to change the test cassette. About 2 seconds later, the meter turns itself off automatically.



The steps that follow are identical to steps 2 to 4 in Chapter 10, section *Replacing the test cassette*, which are accompanied by illustrations (see page 105).

Turn the meter over so that the back is facing upwards.

Push the slide button for the cassette compartment cover upwards so that the cassette compartment cover springs open.

Remove the test cassette from the cassette compartment.



Make sure that no liquid enters the meter. Avoid scratching the measuring optics.

4



Check whether you can see any dirt inside the tip of the cassette.
If so, remove this carefully.

If the measuring optics are visibly soiled, clean the inside of the meter.

5




Carefully dab the measuring optics and the front opening of the meter with a slightly dampened cotton swab.

6




Leave the cassette compartment cover open and close the tip cover.




Carefully dab the interior of the tip cover with a slightly dampened cotton swab.

Remove the dirt that was left behind.


Allow the meter to dry thoroughly.



Open the tip cover again.




Re-insert the test cassette in the meter.



Press the cassette compartment cover closed.

The cassette compartment cover must close with an audible **CLICK**.



Close the tip cover.

Cleaning and disinfecting the finger pricker

To prevent the transmission of infections, you must regularly clean and disinfect the finger pricker and cap:

- at least once a week,
- when there is blood on them,
- always before someone else handles the finger pricker, for example, to assist you.

You need lint-free cloths in which you can wrap the whole finger pricker, water and 70 % isopropanol. Before you disinfect the finger pricker and the cap, you first have to clean them to remove blood and other dirt.



- Visually impaired people must be assisted by a sighted person when cleaning the finger pricker.
- Do not immerse the finger pricker and the cap in the fluids.
- Use only water for cleaning and only 70 % isopropanol for disinfecting.

Preparation

1

Remove the finger pricker from the meter.

2

Remove the cap from the finger pricker.

3

Remove the lancet drum from the finger pricker.

Cleaning

1

Slightly dampen a cloth with water.

2

Wipe the finger pricker and the cap thoroughly; the finger pricker from the outside only and the cap from the outside and the inside.

3

Dry the finger pricker and cap with a dry cloth.

Disinfecting

1

Dampen a cloth well with 70 % isopropanol.

2

Wipe the finger pricker and the cap thoroughly for 2 minutes each; the finger pricker from the outside only and the cap from the outside and the inside.

Place the cap back on again.

3

Wrap the whole finger pricker in the cloth.


Leave the finger pricker wrapped up for 8 minutes.

Unwrap the finger pricker and let it dry in the air.

Changing the batteries

The message *Batteries low Replace batteries* appears on the screen when the batteries are almost empty. It appears every time you turn the meter off if you have not yet changed the batteries. After the message appears for the first time, you can still perform about 50 tests.

Change battery message



Change the batteries as soon as possible.

You need 2 alkaline-manganese batteries type AAA, LR 03, AM 4 or micro (1.5 V) or 2 lithium batteries type AAA, FR 03, L 92 or 2 rechargeable NiMH batteries (type AAA). Always replace both batteries at the same time.

With the batteries supplied, you can perform approximately 500 tests or perform tests for about 1 year. When you use new batteries, the number of tests can vary depending on the battery manufacturer. In order to be able to continue performing a high number of tests, use batteries with as high performance characteristics as possible (high energy content and low self-discharge).

Specific settings increase the power consumption and thereby reduce the life expectancy of the batteries. The following settings increase the power consumption:

- Brightness of screen set to level 3
- Volume set to level 4 or 5
- Acoustic mode on

When you use rechargeable batteries, note the following:

- The number of tests that can be performed per battery charge is much smaller than with regular batteries.
- When the prompt to change the batteries is displayed for the first time, considerably fewer than 50 tests can be performed; with old and used up batteries, it is possible that no more tests can be performed.

Results and settings remain saved in the meter when you change the batteries or if you store the meter without batteries.

The meter has an built-in backup battery. This provides the power to maintain the clock function when there are no batteries inserted. The backup battery has a life expectancy of about 2 years. It cannot be replaced with a new battery.

If the backup battery is empty, the meter loses the set time and date and dashes appear on the screen instead of the time and date. In this case, set the time and date again.



- Visually impaired people must be assisted by a sighted person when changing the batteries.
- Never throw batteries into a fire. They may explode.



- Only remove the batteries when the meter is turned off.
- Remove the batteries if you will not be using the meter for a longer period of time.



Dispose of used batteries in an environmentally-friendly way at an appropriate collection depot.

For information about correct disposal, contact your local council or authority.




Push the battery door in the direction of the arrow as far as it will go to open it.



Lift up the battery door.



Turn the meter so that the batteries fall downwards out of the battery compartment.



Insert two new batteries in the battery compartment. Ensure that the polarity (+ and -) of the batteries is correct.



Close the battery door again.



Push the battery door in the direction of the arrow as far as it will go to close it.

Testing and storage conditions

Temperature

Make sure that the following conditions are met so that the meter and finger pricker operate reliably and you obtain accurate test results:

Storage	Temperature
Blood glucose monitoring system without batteries, without test cassette	-25 to +70 °C
Blood glucose monitoring system with batteries, without test cassette	-10 to +50 °C
Blood glucose monitoring system with batteries, with test cassette	+2 to +30 °C



At temperatures above +50 °C, the batteries could leak and damage the meter.

For blood glucose tests and control tests, the permitted temperature range is between +10 and +40 °C. If the temperature is between +8 and +10 °C or between +40 and +42 °C, the meter will still allow you to perform a test. However, the message *Temperature too low* or *Temperature too high* appears (see *Messages and problems* page 122):



Do not use blood glucose results obtained despite this warning as a basis for making therapeutic decisions. These test results may be incorrect. Incorrect test results can cause the wrong therapy recommendation to be made and thus produce serious adverse health effects.

Tests cannot be performed at temperatures below +8 or above +42 °C. In this case, the E-8 message *Temperature too low: Retest in a warmer location* or *Temperature too high: Retest in a cooler location* appears (see *Messages and problems* page 128).



Never try to artificially speed up a temperature change in your meter, for example, by placing it in the refrigerator or on a radiator. Doing so can damage the meter and cause it to produce incorrect test results. Incorrect test results can cause the wrong therapy recommendation to be made and thus produce serious adverse health effects.

Do not store the finger pricker at very low or high temperatures, for example, in a hot car.

Humidity

Only perform blood glucose tests at a relative humidity of between 15 and 85 %.

Store the meter in a place with a relative humidity of between 15 and 93 %.



Sudden changes in temperature cause condensation to form in or on the meter. If this happens, do not turn the meter on. Allow the meter to return to ambient temperature slowly. Do not store the meter in high moisture areas (for example, in a bathroom).

Light conditions

Displayed text, numbers and symbols appear in yellow. Bright light shining on the display may make them difficult to read. Shield the meter if necessary, with your body, for example.



Keep the meter away from very strong light sources (for example, direct sunlight). These may interfere with the proper functioning of the meter and lead to error messages.

Sources of interference in the surrounding area



- Strong electromagnetic fields may interfere with the proper operation of the meter. Do not use the meter close to sources of strong electromagnetic radiation.
- To avoid electrostatic discharge, do not use the meter in a very dry environment, especially one in which synthetic materials are present.

Messages and problems

Messages

Validity



This message appears at the start of a test if the test cassette is only valid for 10 more days. The message is repeated when the test cassette is only valid for 5, 2 and 1 more day(s).

Press  to start a test.

Changing the battery



This message appears when the batteries are almost empty. It appears every time you turn the meter off if you have not yet changed the batteries. After the message appears for the first time, you can still perform about 50 tests.


[Change the batteries as soon as possible.](#)

Tip cover



This prompt appears if you turn the meter off and the tip cover is still open.

[Close the tip cover.](#)



This message appears if you are in the process of changing a setting (for example, the date) or entering a setting again (for example, a reminder) and open the tip cover.


Close the tip cover again if you want to continue with the setting.

Or

Use \blacktriangledown \blacktriangle to select whether you want to *Continue with settings* or *Test*.

Press \blacksquare .

- *Continue with settings*: The meter returns to the *Settings* menu.
- *Test*: The setting procedure is interrupted (any changes you have already made are lost) and a test area is advanced.



This message appears if you close the tip cover once the meter has advanced a new test area.


Open the tip cover again if you want to continue with the test.

Or

Use \blacktriangledown \blacktriangle to select *Do not test* or *Test*.

Press \blacksquare .

- *Do not test*: The meter opens the Main menu. A test area is lost.
- *Test*: The meter prompts you to open the tip cover.



If you have set reminders, this message appears at the set time.

If you want to perform a test, use \blacktriangledown \blacktriangle to select *Test*.


Or

If you want to be reminded again in 15 minutes, use \blacktriangledown \blacktriangle to select *Snooze*.

Or

Use \blacktriangledown \blacktriangle to select *Cancel* to turn the meter off.

Press \blacksquare .



Do not use blood glucose results obtained despite this warning as a basis for making therapeutic decisions. These test results may be incorrect. Incorrect test results can cause the wrong therapy recommendation to be made and thus produce serious adverse health effects. Move to a place where the temperature is between +10 and +40 °C and wait for the temperature of the meter to adjust to this temperature.

This message appears at the start of a test if the ambient temperature or the temperature of the meter is between +8 and +10 °C **1** or between +40 and +42 °C **2**.

Use **▼▲** to select whether you want to *Test* or *Quit*.

Press **⏻**.

- *Test*: The meter advances a test area.
- *Quit*: The meter opens the Main menu. Close the tip cover.

If you decide to perform the test in spite of the warning, the test result is flagged with the symbol  and stored.


Problems

Problem and possible causes	Solution to the problem
The meter will not turn on.	
The batteries are almost empty, empty or none are inserted.	Insert new batteries.
The batteries were inserted the wrong way.	Remove the batteries and re-insert them as shown in the battery compartment.
The ambient temperature is low and the batteries are almost empty.	Move to a place where the temperature is between +10 and +40 °C and wait for the temperature of the meter to adjust to this temperature. Insert new batteries.
Condensation has made the electronics damp.	Allow the meter time to dry slowly.
The meter is defective.	Contact the customer support and service centre (see page 134).
The meter displays dashes instead of the time and date.	
The meter has no longer saved the time and date. The meter was stored without batteries and the built-in backup battery is empty.	Reset the time and date.
The meter is turned on, but the display remains blank.	
The display is defective.	Contact the customer support and service centre (see page 134).

Error messages


On the following pages you can find a description of all error messages and how to solve them.

- If an error message appears, follow the instructions on the screen.
- If error messages occur frequently, contact the customer support and service centre (see page 134).
- If your meter has been dropped, this can also lead to implausible test results or error messages. In this case, contact the customer support and service centre (see page 134).




The use by period or use by date of the test cassette is exceeded.

Insert a new test cassette in the meter.



This message appears if you start a test and all test areas in the test cassette have been used up.

Insert a new test cassette in the meter.



Remove the test cassette from the meter and re-insert it.

Or

Insert a new test cassette in the meter.

E-1



Cassette
invalid:
Insert
valid
cassette

(E-1)

Insert a test cassette suitable for this meter type.

E-2



Cassette?
Insert new
cassette

(E-2)

There is no test cassette in the meter.

Insert a new test cassette in the meter.

E-3



Meter error:
Retest

(E-3)

Your blood glucose value may be very low.

If a very low blood glucose value matches how you feel, follow the relevant instructions given by your healthcare professional immediately. Repeat the test afterwards.

A meter error has occurred or the meter is defective.

If the error message is still displayed after turning the meter on again, contact the customer support and service centre (see page 134). The meter may be defective.

E-4

Blood applied
incorrectly:
see User's
Manual

(E-4)

You have applied too little blood to the test area.

OR

You did not apply the blood to the centre of the test area or applied it too much to one side.

Repeat the test with a larger amount of blood. Apply the blood drop to the centre of the test area.

E-4

Cassette
dirty:
Clean per
user's manual

(E-4)

Continue

Remove the dirt inside the tip of the cassette and, if necessary, carefully clean the measuring optics.

Press  if you want to go directly to the *Clean* menu.

E-5



Handling
error:
see User's
Manual

(E-5)

E-6



Too bright:
Retest
in shade

(E-6)

You...

- have not washed your fingers or they are soiled or sticky.
- applied a contaminated blood drop (for example, due to food or drink residue on your finger).
- pressed your finger on the test area.
- moved the tape of the test cassette.
- smeared the blood on the test area.
- were trembling or shaking.
- did not remove your finger from the test area as soon as you heard the beep tone.

Wash your hands with warm water and soap and rinse well. Dry your hands thoroughly with a clean towel before obtaining blood.

Touch the test area gently and steadily with the blood drop or control solution. Apply the blood drop or control solution to the centre of the test area. Avoid trembling or shaky movements.

Remove your finger from the test area as soon as the beep tone sounds and *Test in progress* is displayed. The beep tone helps you to obtain a reliable test result.

Move into the shade or shield the meter, for example, with your body.

Repeat the test.

See the next page for the
E-6 error message ►

E-6

Drop applied
too early:
Retest

(E-6)

Do not apply blood or control solution to the test area until the prompt *Apply drop* is displayed.

Repeat the test.

E-7

Meter failure:
Contact
customer
service

(E-7)

If the error message reappears after turning the meter on again:

Contact the customer support and service centre (see page 134).

E-8

Temperature
too low:
Retest
in a warmer
location

(E-8)

Move to a place where the temperature is at least +10 °C and wait for the temperature of the meter to adjust to this temperature.

E-8

Temperature
too high:
Retest
in a cooler
location

(E-8)

Move to a place where the temperature is at most +40 °C and wait for the temperature of the meter to adjust to this temperature.

E-9

Batteries
drained:
Replace
batteries

(E-9)

Insert two new batteries.

Discarding the blood glucose monitoring system

During blood glucose testing, the meter may come into contact with blood. Used meters therefore carry a risk of infection. Discard your used meter, after removing the batteries, according to local regulations.

The meter falls outside the scope of the European Directive 2012/19/EC (Directive on waste electrical and electronic equipment).


✕ Dispose of used batteries in an environmentally-friendly way at an appropriate collection depot. For information about correct disposal, contact your local council or authority.

Used lancet drums and test cassettes can be disposed of in household waste if no other regulations apply locally.

Removing the backup battery

In addition to the two batteries in the battery compartment, there is a third battery in the side of the battery compartment; the backup battery. You must also remove this battery before discarding the meter.

Remove the backup battery as follows:



Use a small screwdriver to break out the partition wall in front of the battery.

The battery then drops into the battery compartment.



Turn the meter over so that the backup battery falls out.

Technical data

Meter type	Accu-Chek Mobile Model U1
Catalogue no./serial no.	See type plate on the back of the meter
Test principle	Determination of glucose in fresh capillary blood by reflectance photometry. When using different samples, refer to the package insert of the Accu-Chek Mobile test cassette. Your meter displays blood glucose values that refer to plasma although you always apply whole blood to the test area. You can find information on how the test works and on the referenced test principle in the package insert of the Accu-Chek Mobile test cassette.
Measuring interval	See the package insert of the Accu-Chek Mobile test cassette
Blood volume	See the package insert of the Accu-Chek Mobile test cassette
Measuring time	Approx. 5 seconds (depending on the concentration)
Power supply	2 alkaline-manganese batteries (1.5 V; type AAA, LR 03, AM 4 or micro) or 2 lithium batteries (1.5 V; type AAA, FR 03, L 92), backup battery: 3-volt lithium coin cell type CR1025
Battery life	Approximately 500 tests or approximately 1 year (less if the brightness of the screen is set to level 3 or the volume is set to level 4 or 5, or in acoustic mode due to the higher power consumption)
Automatic power off	After 1 or 2 minute(s), depending on the operating status

Temperature	
During testing	+10 to +40 °C
During storage	Meter without batteries and without test cassette: -25 to +70 °C
	Meter with batteries and without test cassette: -10 to +50 °C
	Meter with batteries and with test cassette: +2 to +30 °C
Humidity	
During testing	15 to 85 % relative humidity
During storage	15 to 93 % relative humidity
Altitude	Sea level to 4,000 m above sea level
Memory capacity	2,000 results with time and date, averages for 7, 14, 30 and 90 days
Dimensions	121 × 63 × 20 mm with finger pricker
Weight	Approx. 129 g with finger pricker, batteries, test cassette and lancet drum
Display	OLED (Organic Light Emitting Diode) display
Interface	USB (Micro B)
Electromagnetic compatibility	<p>This device meets the electromagnetic immunity requirements as per EN ISO 15197 Annex A. The chosen basis for immunity testing (on electrostatic discharge) was basic standard IEC 61000-4-2.</p> <p>In addition, the device meets the electromagnetic emissions requirements as per EN 61326. Its electromagnetic emission is thus low. Interference from the device to other electrically-driven equipment is not anticipated.</p>

Performance assessment	The performance characteristics of the Accu-Chek Mobile system (Accu-Chek Mobile meter and Accu-Chek Mobile test cassette) were determined using capillary blood from patients with diabetes (system accuracy), venous blood (repeatability) and control solution (intermediate precision).
Calibration and traceability	The system is calibrated with whole blood containing various glucose concentrations as a calibrator. The reference values are obtained using the hexokinase method which is calibrated using the ID-GCMS method. The reference method is traceable to a NIST standard using the ID-GCMS method, which is the method of highest metrological quality (order).

The Accu-Chek Mobile blood glucose monitoring system complies with the requirements of EN ISO 15197.

Declaration of conformity

The Accu-Chek Mobile blood glucose meter conforms with the basic requirements and other relevant regulations of the European Directive 1999/5/EC on radio and telecommunications terminal equipment (R&TTE). The conformity declaration may be found at the following Internet address:

<http://declarations.accu-chek.com>

System components

Accu-Chek Mobile meter (Model U1)

Accu-Chek Mobile test cassettes

Use only these test cassettes when you perform blood glucose tests using the Accu-Chek Mobile meter.

Accu-Chek Mobile control solutions

Use only these control solutions when you perform control tests using the Accu-Chek Mobile meter and the respective test cassette.



Contact your customer support and service centre (see page 134) for information on where you can obtain the test cassettes and control solutions.

Accu-Chek FastClix finger pricker

Accu-Chek FastClix lancet drums

Use only these lancet drums when you obtain blood with the Accu-Chek FastClix M1 finger pricker. You can find the type designation **1** on the side of the finger pricker.



Customer Support and Service Centre

If you need advice on how to operate the Accu-Chek Mobile meter or the Accu-Chek FastClix finger pricker, if you seem to be obtaining implausible test results, or if you suspect that the meter, test cassette, finger pricker or lancet drum might be defective, contact the Customer Support and Service Centre. Do not attempt to repair or modify the meter or finger pricker yourself. Our staff will help solve any problems you might be experiencing with the meter, test cassette, finger pricker or lancet drum from Roche.

Australia

Accu-Chek Enquiry Line: 1800 251 816
Pump Support: 1800 633 457
www.accu-chek.com.au

Singapore

Accu-Chek ExtraCare line: 6272 9200
www.accu-chek.com.sg

United Kingdom

Roche Diabetes Care Limited
Charles Avenue, Burgess Hill
West Sussex, RH15 9RY, United Kingdom
Accu-Chek Customer Careline ¹⁾

UK Freephone number: 0800 701 000

ROI Freephone number: 1 800 709 600

¹⁾ calls may be recorded for training purposes

www.accu-chek.co.uk

www.accu-chek.ie

Index

A

acoustic mode 49, 50
after meal (flag) 38, 39
apple core (symbol) 38
apple (symbol) 38, 69
applicator (symbol) 38
apply blood 28, 34
asterisk (symbol) 38
available tests. *See* tests
averages 67

B

battery
 changing 114
 discarding 115, 129
 life expectancy 114, 115, 120, 130
 symbol 35
 type 114, 130
battery door 8, 116
beep tones 49
 test result announcement
 (acoustic mode) 91
 turning on/off 50
before meal (flag) 38, 39
blood drop, well-formed 28
blood glucose monitoring system
 cleaning 107
brightness, screen 63
button
 down ~ 9, 13, 65
 power/enter ~ 9, 11
 up ~ 9, 13, 103

C

checking, meter 94
cleaning
 finger pricker 112
 meter 107
concentration range 98, 99
connecting, to computer (PC) 73
Continua Health Alliance 71
control solution 94
control test 94
 performing 95
 preparing 95
control test. *Siehe* control test
customer support and service
 centre 134

D

data transfer 71, 74
default
 data transfer 71
 reports 71
discarding
 batteries 115, 129
 meter 129
 meter, test cassette 129
disinfecting
 finger pricker 112
display 8, 131
display check 19, 101
down button 9

E

electromagnetic compatibility 131
enter button 9
error messages 124
 in acoustic mode 93

F

finger pricker
 cleaning and disinfecting 112
 docking 26
 obtaining blood with the ~ 33
 overview 8
 undocking 26
flag
 after meal 39
 before meal 39
 control test 39, 97
 other 39

H

hands
 dry 28
 wash 28
HI (screen) 36, 42
humidity 118

I

intended use 3

L

lancet drum 8, 22
 discarding 25
 inserting 22
 replacing 25
list report 84
loading a lancet 24
LO (screen) 36, 42

M

making settings 45
measurement conditions 117
measuring range 36, 130
measuring time 34, 130

memory 64, 131
memory, results 64
menu
 description of menu structure
 10
 open 11
meter
 checking 94
 cleaning 107
 discarding 129
 overview 7
 storing. *See* storage conditions
 turning off. *See* turning off
 turning on. *See* turning on
micro B connector 70

N

number of lancets 8
number of tests 11, 103

O

one-time reminder 41
overview
 finger pricker 8
 meter 7
 setting options 45

P

PC analysis 71
penetration depth
 indicator 8, 23
 setting 23
performance assessment 132
performing a blood glucose test
 27
power button 8, 9
problems 123

R

release button 8, 33
reminder 121
 changing 54
 deleting 54
 setting 41, 51, 52
 turning on/off 53
reports
 displaying 75
 printing 80
retrieving results from memory 65

S

screen 10
 setting the brightness 63
 symbols shown on the ~ 35
setting flag, result 37
setting the date 57
setting the time 55
setting the time format 59
setting tones 49
slide button
 cassette compartment cover 8,
 16, 105, 109
 undocking the finger pricker 8,
 26
sources of error
 control test 99
 test 43
sources of interference 119
standard day report 82
standard week report 83
storage conditions 117
symbol applicator 38
symbols 138
 during testing 35
system components 133

T

target range
 changing 62
 setting 60
 symbols 35, 60
 turning on/off 61
technical data 130
temperature 117, 131
test
 performing 27
 sources of error 43
test area 8
test cassette 8
 discarding 106, 129
 inserting first ~ 15
 replacing 103
 use by date 18
 validity 102, 120
test principle 130
test result
 averages 67
 flagging 37
 retrieving from memory 65
test results
 transferring 74
tests (test areas)
 display of available ~ 11, 20,
 102
thermometer (symbol) 35, 122
tip cover 8
transferring, test results 74
trend report 81
turning off, meter 9, 11
turning on, meter 11
 connecting, to PC 73
 retrieving results from memory
 65
 starting a test 30

type plate 8, 14

U

unit of measurement (mg/dL,
mmol/L) 14

up button 9

USB A connector 70

USB port 8, 70, 73

use by date

control solution 100

use by period, test cassette 18,
124

V

validity, test cassette 18, 124

displaying 102

Explanation of symbols

The following symbols may appear on the packaging, on the type plate of the meter and on the finger pricker. They have the following meanings:



Consult instructions for use



Caution, refer to safety-related notes in the instructions for use accompanying this product.



Temperature limitation (store at)



Use by



Use only once



Sterilized using irradiation



Manufacturer



Catalogue number



Batch code



In vitro diagnostic medical device



Global Trade Item Number



Blood glucose meter: This product fulfils the requirements of the European Directive 1999/5/EC on radio and telecommunications terminal equipment (R&TTE).



Blood glucose meter: This product fulfils the requirements of the European Directive 98/79/EC on in vitro diagnostic medical devices.



Finger pricker and lancet drum:
These products fulfil the requirements of the European Directive 93/42/EEC on medical devices.



This device complies with Part 15 of the FCC Rules and with RSS-210 of Industry Canada.



The compliance mark indicates that the product complies with the applicable standard and establishes a traceable link between the equipment and the manufacturer, importer or their agent responsible for compliance and for placing it on the Australian and New Zealand market.

The explanation of other symbols can be found in the instructions for use and inserts accompanying components within the packaging.



According to Continua Health Alliance guidelines



Certified by USB implementers forum

Appendix

Low blood glucose index or high blood glucose index

These figures represent the frequency and the resulting risk of blood glucose values being too low or too high. Figures should be as low as possible.

The following table provides an overview to assess the risk of blood glucose values being too low or too high:

Risk	Low blood glucose index	High blood glucose index
minimal	≤ 1.1	≤ 5.0
low	1.1–2.5	5.0–10.0
medium	2.5–5.0	10.0–15.0
high	> 5.0	> 15.0



The index values for *low blood glucose* or *high blood glucose* in the table are **not** blood glucose values. Ask your healthcare professional if you want to change your therapy based on the index values.

References

Low Blood Glucose Index / High Blood Glucose Index

Boris P. Kovatchev, Martin Straume, Daniel J. Cox, Leon S. Farhy (2001)

“Risk analysis of blood glucose data: a quantitative approach to optimizing the control of insulin dependent diabetes.”

Journal of Theoretical Medicine, **3**: pp 1-10.

Boris P. Kovatchev, Daniel J. Cox, Anand Kumar, Linda Gonder-Frederick, William L. Clarke (2003)

“Algorithmic Evaluation of Metabolic Control and Risk of Severe Hypoglycemia in Type 1 and Type 2 Diabetes Using Self-Monitoring Blood Glucose Data”

Diabetes Technology & Therapeutics, **5**(5): pp 817-828

Boris P. Kovatchev (2006)

“Is Glycemic Variability Important to Assessing Antidiabetes Therapies?”

Current Diabetes Reports, **6**: pp 350-356

ACCU-CHEK® 360° reports License

By using this software, you, or the entity on whose behalf you are using the software (here in after “Licensee”) agree to be legally bound to the following provisions.

1 License

- 1.1 Roche Diagnostics GmbH, Sandhofer Strasse 116, 68305 Mannheim, Germany (“Roche”) grants to Licensee, a non-exclusive, perpetual and worldwide license to use the ACCU-CHEK 360° reports software, here in after referred to as SOFTWARE, upon the terms and conditions contained in this Software License Agreement.
- 1.2 Licensee may use the SOFTWARE on all computer systems owned, leased or otherwise controlled by Licensee. Licensee may use the SOFTWARE only to manage the ACCU-CHEK 360° reports.
- 1.3 Due to Roche being obliged to commit market observations, Licensee shall notify Roche of name and address of the receiving party in case of any transfer of the SOFTWARE product and shall oblige the receiving party to enter into an equivalent agreement with Roche.
- 1.4 Except as expressly permitted by Swiss law, Licensee may not: Copy the SOFTWARE in whole or in part except as expressly provided in this Software License Agreement or for purposes of backup; rent, license or sublicense the SOFTWARE; create derivative works based on SOFTWARE; modify, adapt, translate, reverse engineer, decompile or disassemble the SOFTWARE, separate the SOFTWARE into its component parts, or in any way reverse engineer or attempt to reconstruct or discover any source code or algorithms of the SOFTWARE by any means whatsoever unless and until Licensee has first requested the required information from Roche in writing, and Roche, at its sole discretion, has not complied with Licensee’s request within a commercially reasonable period of time; remove any product identification, trademark, copyright, confidentiality, proprietary or other notices contained on or within the SOFTWARE.
- 1.5 If Licensee is an entity, then Licensee shall cause Licensee’s employees, if any, to comply with the terms and conditions of the Software License Agreement.
- 1.6 Licensee shall cooperate with Roche, and shall render all reasonable assistance requested by Roche, to assist Roche in preventing and identifying any use of, or access to, the SOFTWARE in violation of the Software License Agreement.
- 1.7 Licensee acknowledges and agrees to the terms and conditions of any open source software license contained in or referred to in the software including relevant copyright notices.

2 Proprietary Rights

All right, title and interest including, but not limited to, copyright, trademarks and other intellectual property rights in and to the SOFTWARE are owned by Roche and its licensors, if any. Such rights are protected by the Urheberrechtsgesetz (URG), Markenschutzgesetz (MSchG) and other laws. Roche retains all rights not expressly granted herein.

3 Warranty

- 3.1 Roche warrants that the SOFTWARE licensed under this Software License Agreement will substantially provide the key functions described in its product specification in effect on the date of conclusion of this Software License Agreement and that the SOFTWARE will be free of defects for a period of one year from shipment.
- 3.2 During the warranty period Roche shall, at Roche's sole discretion correct errors detected in licensed software or provide a reasonable workaround SOFTWARE, after receiving notification from Licensee of such errors. Error corrections will if technically possible be provided by way of a patch or update to the SOFTWARE. The warranties set out herein are exclusive of and in lieu of all other conditions and warranties, either express or implied, statutory or otherwise. Notwithstanding this, Roche may from time to time provide proactively updates (including service releases).
- 3.3 This warranty is void if failure is the result of accident, abuse, misuse, alteration, use with unauthorized software or hardware or not having installed latest updates of the SOFTWARE provided by Roche according to Sec. 3.2.

4 Limitation of Liability

- 4.1 Irrespective of the cause of action, Roche liability shall be limited to damages caused by Roche, its employees or subcontractors due to wilful misconduct, gross negligence or, in case of a violation of an essential contractual duty, simple negligence.
- 4.2 If Roche is held liable for simple negligence under Section 4.1, Roche's liability shall be limited to those typical damages that were reasonably foreseeable at the time the Agreement was concluded or, at the latest at the time of the violation of the relevant contractual duty.
- 4.3 Roche's liability for any damages caused by the breach of a specific warranty (Garantie) or for damages to be compensated under the Product Liability Act and for damages due to loss of life, injury or prejudice to health remains unaffected.
- 4.4 The Licensee is obliged to carry out regular backups. In case of any loss of data Roche is only liable for the costs to restore such backups.

5 Miscellaneous

- 5.1 Changes to this Software License Agreement require written form to be effective. This also applies for any amendment or rescinding of this written form clause.
- 5.2 This Software License Agreement is subject to Swiss law, to the exclusion of the UN Convention on the International Sale of Goods as well as to the exclusion of the regulations of international civil law that are alterable subject to mutual agreement.
- 5.3 For disputes arising from or in connection with this contract, the courts of Basel, Switzerland have jurisdiction internationally and locally.

- 5.4 Should individual clauses of the Software License Agreement be ineffective, the effectiveness of the contract remains unaffected otherwise. Ineffective clauses shall be replaced by the legal regulation.